

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035348.D
 Acq On : 27 Apr 2023 17:54
 Operator : JC/MD
 Sample : 02495-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GMW-5

Quant Time: Apr 28 05:08:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	231525	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	382487	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	343993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	157240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	137690	49.294	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.580%
35) Dibromofluoromethane	5.385	113	115464	49.184	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.360%
50) Toluene-d8	8.647	98	445558	50.312	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.620%
62) 4-Bromofluorobenzene	11.079	95	164499	49.378	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.760%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	11952	9.852	ug/l	96
44) Trichloroethene	7.129	130	2231	0.770	ug/l	75
95) Naphthalene	13.780	128	2247	0.214	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035348.D
Acq On : 27 Apr 2023 17:54
Operator : JC/MD
Sample : 02495-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GMW-5

Quant Time: Apr 28 05:08:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

