

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035349.D
 Acq On : 27 Apr 2023 18:17
 Operator : JC/MD
 Sample : 02495-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Apr 28 05:08:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	224211	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	369366	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	342746	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159589	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	135858	50.225	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.460%	
35) Dibromofluoromethane	5.385	113	113519	50.073	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.140%	
50) Toluene-d8	8.646	98	435186	50.886	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.780%	
62) 4-Bromofluorobenzene	11.079	95	165342	51.394	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.780%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2819	2.400	ug/l	96
44) Trichloroethene	7.128	130	1679	0.600	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035349.D
Acq On : 27 Apr 2023 18:17
Operator : JC/MD
Sample : 02495-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 28 05:08:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

