

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035351.D
 Acq On : 27 Apr 2023 19:03
 Operator : JC/MD
 Sample : 02494-01 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Apr 28 05:09:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	226530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	369676	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	335982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135531	49.591	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.180%
35) Dibromofluoromethane	5.385	113	115055	50.708	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.420%
50) Toluene-d8	8.646	98	433500	50.647	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.300%
62) 4-Bromofluorobenzene	11.079	95	162814	50.566	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.140%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.373	62	103731	47.431	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	271581	96.689	ug/l	85
44) Trichloroethene	7.141	130	1265	0.452	ug/l	81
93) 1,2,4-Trichlorobenzene	13.591	180	718	0.212	ug/l	89
94) Hexachlorobutadiene	13.725	225	650	0.437	ug/l	92
96) 1,2,3-Trichlorobenzene	13.963	180	749	0.219	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035351.D
Acq On : 27 Apr 2023 19:03
Operator : JC/MD
Sample : 02494-01 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Apr 28 05:09:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

