

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035352.D
 Acq On : 27 Apr 2023 19:27
 Operator : JC/MD
 Sample : 02492-01DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-04IDL

Quant Time: Apr 28 05:10:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	231325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	377485	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	344257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140080	50.193	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.380%	
35) Dibromofluoromethane	5.391	113	116179	50.144	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.280%	
50) Toluene-d8	8.647	98	443641	50.759	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.520%	
62) 4-Bromofluorobenzene	11.079	95	170678	51.912	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.820%	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.489	96	53661	18.708	ug/l	86
44) Trichloroethene	7.123	130	233484	81.668	ug/l	97
64) Tetrachloroethene	9.281	164	625	0.237	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035352.D
Acq On : 27 Apr 2023 19:27
Operator : JC/MD
Sample : 02492-01DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-04IDL

Quant Time: Apr 28 05:10:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

