

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035353.D
 Acq On : 27 Apr 2023 19:50
 Operator : JC/MD
 Sample : 02495-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-01

Quant Time: Apr 28 05:10:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	225496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	367679	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	327989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	134969	49.612	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.220%
35) Dibromofluoromethane	5.385	113	113470	50.281	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.560%
50) Toluene-d8	8.647	98	428157	50.294	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.580%
62) 4-Bromofluorobenzene	11.079	95	154065	48.108	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.220%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20886	17.677	ug/l	100
27) cis-1,2-Dichloroethene	4.501	96	607	0.217	ug/l	69
44) Trichloroethene	7.129	130	1379	0.495	ug/l	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035353.D
Acq On : 27 Apr 2023 19:50
Operator : JC/MD
Sample : 02495-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-01

Quant Time: Apr 28 05:10:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

