

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035361.D
 Acq On : 28 Apr 2023 03:26
 Operator : JC/MD
 Sample : 02522-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-13D

Quant Time: Apr 28 05:43:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	234789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	380026	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	343478	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152328	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138513	48.900	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.800%
35) Dibromofluoromethane	5.385	113	117961	50.573	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.140%
50) Toluene-d8	8.647	98	449798	51.120	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.240%
62) 4-Bromofluorobenzene	11.079	95	161575	48.814	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.620%
Target Compounds						
					Qvalue	
30) Chloroform	5.099	83	1955	0.433	ug/l	96
44) Trichloroethene	7.135	130	647	0.225	ug/l	54
64) Tetrachloroethene	9.275	164	4720	1.795	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035361.D
Acq On : 28 Apr 2023 03:26
Operator : JC/MD
Sample : 02522-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-13D

Quant Time: Apr 28 05:43:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

