

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035366.D
 Acq On : 28 Apr 2023 05:22
 Operator : JC/MD
 Sample : 02522-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Apr 28 06:44:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	232689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	380732	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	347472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	164402	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	137271	48.899	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.800%
35) Dibromofluoromethane	5.385	113	116910	50.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	8.647	98	447533	50.768	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.540%
62) 4-Bromofluorobenzene	11.079	95	170396	51.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.760%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2715	2.227	ug/l	89
19) Methyl tert-butyl Ether	3.111	73	8146	1.051	ug/l #	86
40) Benzene	6.043	78	2672	0.255	ug/l #	78
52) Toluene	8.726	92	1640	0.247	ug/l	89
64) Tetrachloroethene	9.275	164	4023	1.512	ug/l	94
67) Ethyl Benzene	10.195	91	6297	0.484	ug/l #	88
68) m/p-Xylenes	10.299	106	14076	2.822	ug/l	96
69) o-Xylene	10.640	106	6950	1.421	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	3682	0.363	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	11796	1.159	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035366.D
Acq On : 28 Apr 2023 05:22
Operator : JC/MD
Sample : 02522-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Apr 28 06:44:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

