

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035377.D
 Acq On : 28 Apr 2023 09:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 28 11:42:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.529	0.545	-3.0	80	0.00
3 P	Chloromethane	0.481	0.438	8.9	75	0.00
4 C	Vinyl Chloride	0.483	0.467	3.3#	79	0.00
5 T	Bromomethane	0.321	0.292	9.0	75	0.00
6 T	Chloroethane	0.282	0.272	3.5	80	0.00
7 T	Trichlorofluoromethane	0.755	0.776	-2.8	86	0.00
8 T	Diethyl Ether	0.275	0.268	2.5	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.497	1.4	82	0.00
10 T	Methyl Iodide	0.636	0.576	9.4	70	0.00
11 T	Tert butyl alcohol	0.110	0.111	-0.9	79	0.00
12 CM	1,1-Dichloroethene	0.486	0.483	0.6#	83	0.00
13 T	Acrolein	0.108	0.090	16.7	67	0.00
14 T	Allyl chloride	0.855	0.817	4.4	78	0.00
15 T	Acrylonitrile	0.275	0.265	3.6	79	0.00
16 T	Acetone	0.262	0.226	13.7	72	0.00
17 T	Carbon Disulfide	1.164	1.173	-0.8	81	0.00
18 T	Methyl Acetate	1.013	1.007	0.6	80	0.00
19 T	Methyl tert-butyl Ether	1.666	1.717	-3.1	84	0.00
20 T	Methylene Chloride	0.601	0.550	8.5	83	0.00
21 T	trans-1,2-Dichloroethene	0.527	0.525	0.4	83	0.00
22 T	Diisopropyl ether	1.691	1.659	1.9	80	0.00
23 T	Vinyl Acetate	1.162	1.175	-1.1	79	0.00
24 P	1,1-Dichloroethane	0.953	0.973	-2.1	83	0.00
25 T	2-Butanone	0.382	0.350	8.4	75	0.00
26 T	2,2-Dichloropropane	0.746	0.574	23.1	62	0.00
27 T	cis-1,2-Dichloroethene	0.620	0.614	1.0	83	0.00
28 T	Bromochloromethane	0.406	0.353	13.1	69	0.00
29 T	Tetrahydrofuran	0.245	0.227	7.3	76	0.00
30 C	Chloroform	0.962	1.002	-4.2#	85	0.00
31 T	Cyclohexane	0.856	0.828	3.3	79	0.00
32 T	1,1,1-Trichloroethane	0.843	0.890	-5.6	85	0.00
33 S	1,2-Dichloroethane-d4	0.603	0.544	9.8	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.307	0.277	9.8	74	0.00
36 T	1,1-Dichloropropene	0.479	0.463	3.3	84	0.00
37 T	Ethyl Acetate	0.438	0.424	3.2	77	0.00
38 T	Carbon Tetrachloride	0.455	0.482	-5.9	85	0.00
39 T	Methylcyclohexane	0.583	0.542	7.0	77	0.00
40 TM	Benzene	1.375	1.334	3.0	82	0.00
41 T	Methacrylonitrile	0.251	0.249	0.8	80	0.00
42 TM	1,2-Dichloroethane	0.490	0.489	0.2	84	0.00
43 T	Isopropyl Acetate	0.735	0.728	1.0	80	0.00
44 TM	Trichloroethene	0.379	0.376	0.8	83	0.00
45 C	1,2-Dichloropropane	0.347	0.344	0.9#	82	0.00
46 T	Dibromomethane	0.238	0.241	-1.3	85	0.00
47 T	Bromodichloromethane	0.438	0.466	-6.4	85	0.00
48 T	Methyl methacrylate	0.361	0.363	-0.6	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035377.D
 Acq On : 28 Apr 2023 09:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 28 11:42:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	80	0.00
50 S	Toluene-d8	1.158	1.020	11.9	72	0.00
51 T	4-Methyl-2-Pentanone	0.452	0.434	4.0	79	0.00
52 CM	Toluene	0.871	0.859	1.4#	82	0.00
53 T	t-1,3-Dichloropropene	0.458	0.486	-6.1	81	0.00
54 T	cis-1,3-Dichloropropene	0.518	0.528	-1.9	79	0.00
55 T	1,1,2-Trichloroethane	0.347	0.343	1.2	83	0.00
56 T	Ethyl methacrylate	0.517	0.539	-4.3	84	0.00
57 T	1,3-Dichloropropane	0.594	0.584	1.7	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.252	2.7	76	0.00
59 T	2-Hexanone	0.339	0.324	4.4	77	0.00
60 T	Dibromochloromethane	0.333	0.366	-9.9	85	0.00
61 T	1,2-Dibromoethane	0.358	0.365	-2.0	83	0.00
62 S	4-Bromofluorobenzene	0.435	0.385	11.5	73	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.383	0.376	1.8	85	0.00
65 PM	Chlorobenzene	1.082	1.040	3.9	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.376	0.390	-3.7	85	0.00
67 C	Ethyl Benzene	1.873	1.849	1.3#	82	0.00
68 T	m/p-Xylenes	0.718	0.710	1.1	82	0.00
69 T	o-Xylene	0.704	0.700	0.6	83	0.00
70 T	Styrene	1.141	1.171	-2.6	83	0.00
71 P	Bromoform	0.262	0.291	-11.1	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.655	3.758	-2.8	84	0.00
74 T	N-amyl acetate	1.374	1.421	-3.4	79	0.00
75 P	1,1,2,2-Tetrachloroethane	1.134	1.131	0.3	82	0.00
76 T	1,2,3-Trichloropropane	0.997	0.938	5.9	82	0.00
77 T	Bromobenzene	0.909	0.931	-2.4	83	0.00
78 T	n-propylbenzene	4.198	4.198	0.0	81	0.00
79 T	2-Chlorotoluene	2.623	2.601	0.8	83	0.00
80 T	1,3,5-Trimethylbenzene	3.082	3.163	-2.6	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.308	0.292	5.2	73	0.00
82 T	4-Chlorotoluene	2.946	2.959	-0.4	83	0.00
83 T	tert-Butylbenzene	3.092	3.134	-1.4	84	0.00
84 T	1,2,4-Trimethylbenzene	3.097	3.154	-1.8	82	0.00
85 T	sec-Butylbenzene	3.793	3.764	0.8	80	0.00
86 T	p-Isopropyltoluene	3.226	3.225	0.0	80	0.00
87 T	1,3-Dichlorobenzene	1.727	1.715	0.7	82	0.00
88 T	1,4-Dichlorobenzene	1.810	1.720	5.0	82	0.00
89 T	n-Butylbenzene	2.723	2.600	4.5	75	0.00
90 T	Hexachloroethane	0.494	0.523	-5.9	80	0.00
91 T	1,2-Dichlorobenzene	1.727	1.708	1.1	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.244	-7.5	82	0.00
93 T	1,2,4-Trichlorobenzene	1.085	1.080	0.5	82	0.00
94 T	Hexachlorobutadiene	0.476	0.445	6.5	79	0.00
95 T	Naphthalene	3.342	3.531	-5.7	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035377.D
Acq On : 28 Apr 2023 09:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 28 11:42:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.093	1.071	2.0	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6