

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042822\
 Data File : VX028395.D
 Acq On : 28 Apr 2022 12:35
 Operator : JC/MD
 Sample : VX0428WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0428WBL01

Quant Time: Apr 29 04:07:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	411355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	709191	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	670904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	318555	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	291054	54.265	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.520%	
35) Dibromofluoromethane	5.391	113	262442	52.768	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.540%	
50) Toluene-d8	8.653	98	973643	50.970	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.940%	
62) 4-Bromofluorobenzene	11.079	95	349087	47.795	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.580%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042822\
Data File : VX028395.D
Acq On : 28 Apr 2022 12:35
Operator : JC/MD
Sample : VX0428WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBL01

Quant Time: Apr 29 04:07:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

