

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042822\
 Data File : VX028401.D
 Acq On : 28 Apr 2022 15:23
 Operator : JC/MD
 Sample : N2577-03ME
 Misc : 6.86g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GT-06-(8.5-10)ME

Quant Time: Apr 29 04:09:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	579278	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.757	114	916519	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	800007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	449114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	303256	40.150	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	80.300%
35) Dibromofluoromethane	5.385	113	272852	42.451	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	84.900%
50) Toluene-d8	8.647	98	1044662	42.317	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	84.640%
62) 4-Bromofluorobenzene	11.085	95	423158	44.830	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.660%
Target Compounds						
					Qvalue	
25) 2-Butanone	4.587	43	25901	5.790	ug/l	95
39) Methylcyclohexane	7.373	83	53578	5.065	ug/l	94
73) Isopropylbenzene	10.963	105	228301	7.396	ug/l	99
78) n-propylbenzene	11.305	91	538787	15.740	ug/l	98
83) tert-Butylbenzene	11.719	119	41977	1.569	ug/l #	70
84) 1,2,4-Trimethylbenzene	11.756	105	42773	1.642	ug/l	99
85) sec-Butylbenzene	11.896	105	719167	22.278	ug/l #	78
89) n-Butylbenzene	12.335	91	620441	27.132	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042822\
Data File : VX028401.D
Acq On : 28 Apr 2022 15:23
Operator : JC/MD
Sample : N2577-03ME
Misc : 6.86g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GT-06-(8.5-10)ME

Quant Time: Apr 29 04:09:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

