

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
 Data File : VX035382.D
 Acq On : 28 Apr 2023 11:44
 Operator : JC/MD
 Sample : VX0428WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0428WBS01

Quant Time: Apr 28 15:40:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	288171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	446481	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	395148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	195798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	183455	52.768	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.540%	
35) Dibromofluoromethane	5.385	113	149568	54.579	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.160%	
50) Toluene-d8	8.647	98	554818	53.670	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.340%	
62) 4-Bromofluorobenzene	11.079	95	212213	54.570	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.140%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	61623	20.210	ug/l	100
3) Chloromethane	1.294	50	48789	17.589	ug/l	100
4) Vinyl Chloride	1.373	62	51510	18.515	ug/l	98
5) Bromomethane	1.617	94	33456	18.093	ug/l	98
6) Chloroethane	1.691	64	29825	18.330	ug/l	96
7) Trichlorofluoromethane	1.886	101	87719	20.162	ug/l	96
8) Diethyl Ether	2.129	74	30030	18.931	ug/l	92
9) 1,1,2-Trichlorotrifluo...	2.331	101	56599	19.478	ug/l	94
10) Methyl Iodide	2.453	142	58752	16.020	ug/l	98
11) Tert butyl alcohol	3.013	59	61613	97.332	ug/l #	73
12) 1,1-Dichloroethene	2.318	96	53080	18.965	ug/l	97
13) Acrolein	2.239	56	56622	90.954	ug/l	100
14) Allyl chloride	2.666	41	91884	18.655	ug/l	97
15) Acrylonitrile	3.068	53	147637	93.249	ug/l	99
16) Acetone	2.398	43	150428	99.624	ug/l	98
17) Carbon Disulfide	2.514	76	123655	18.430	ug/l	99
18) Methyl Acetate	2.709	43	111712	19.132	ug/l	96
19) Methyl tert-butyl Ether	3.111	73	194002	20.204	ug/l	98
20) Methylene Chloride	2.788	84	61685	17.817	ug/l	95
21) trans-1,2-Dichloroethene	3.093	96	58906	19.378	ug/l	93
22) Diisopropyl ether	3.757	45	191873	19.686	ug/l	99
23) Vinyl Acetate	3.721	43	663951	99.102	ug/l	98
24) 1,1-Dichloroethane	3.611	63	107793	19.634	ug/l	99
25) 2-Butanone	4.568	43	210726	95.614	ug/l	95
26) 2,2-Dichloropropane	4.471	77	86743	20.167	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	69487	19.447	ug/l	95
28) Bromochloromethane	4.897	49	48115	20.565	ug/l #	82
29) Tetrahydrofuran	5.013	42	129931	91.966	ug/l	93
30) Chloroform	5.092	83	112803	20.353	ug/l	99
31) Cyclohexane	5.470	56	95058	19.275	ug/l	95
32) 1,1,1-Trichloroethane	5.385	97	98424	20.264	ug/l	98
36) 1,1-Dichloropropene	5.696	75	83476	19.509	ug/l	98
37) Ethyl Acetate	4.714	43	75762	19.386	ug/l #	94
38) Carbon Tetrachloride	5.678	117	84512	20.805	ug/l	99
39) Methylcyclohexane	7.379	83	101063	19.403	ug/l	98
40) Benzene	6.037	78	243711	19.855	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
 Data File : VX035382.D
 Acq On : 28 Apr 2023 11:44
 Operator : JC/MD
 Sample : VX0428WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0428WBS01

Quant Time: Apr 28 15:40:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	43388	19.375	ug/l	96
42) 1,2-Dichloroethane	6.086	62	88843	20.300	ug/l	96
43) Isopropyl Acetate	6.336	43	127770	19.457	ug/l	94
44) Trichloroethene	7.129	130	68282	20.193	ug/l	90
45) 1,2-Dichloropropane	7.427	63	62505	20.198	ug/l	98
46) Dibromomethane	7.580	93	43200	20.328	ug/l	92
47) Bromodichloromethane	7.824	83	80820	20.641	ug/l	95
48) Methyl methacrylate	7.696	41	64115	19.904	ug/l	93
49) 1,4-Dioxane	7.720	88	32154	409.510	ug/l	93
51) 4-Methyl-2-Pentanone	8.573	43	398427	98.756	ug/l	95
52) Toluene	8.720	92	158294	20.345	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	85001	20.788	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	96512	20.882	ug/l	92
55) 1,1,2-Trichloroethane	9.153	97	62979	20.304	ug/l	97
56) Ethyl methacrylate	9.116	69	95039	20.569	ug/l	90
57) 1,3-Dichloropropane	9.305	76	104355	19.670	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	214504	92.844	ug/l	96
59) 2-Hexanone	9.427	43	304677	100.525	ug/l	95
60) Dibromochloromethane	9.518	129	62473	21.022	ug/l	99
61) 1,2-Dibromoethane	9.610	107	66023	20.648	ug/l	99
64) Tetrachloroethene	9.275	164	60482	19.990	ug/l	98
65) Chlorobenzene	10.079	112	168417	19.700	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	61613	20.708	ug/l	98
67) Ethyl Benzene	10.195	91	300745	20.313	ug/l	97
68) m/p-Xylenes	10.299	106	229975	40.538	ug/l	99
69) o-Xylene	10.640	106	113853	20.466	ug/l	99
70) Styrene	10.652	104	188601	20.910	ug/l	97
71) Bromoform	10.799	173	43263	20.905	ug/l #	98
73) Isopropylbenzene	10.963	105	301130	21.039	ug/l	100
74) N-amyl acetate	10.841	43	110207	20.489	ug/l	94
75) 1,1,2,2-Tetrachloroethane	11.213	83	92250	20.766	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	93959	24.057	ug/l	85
77) Bromobenzene	11.195	156	73671	20.707	ug/l	91
78) n-propylbenzene	11.305	91	334888	20.373	ug/l	99
79) 2-Chlorotoluene	11.366	91	206199	20.074	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	254895	21.121	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	22433	18.602	ug/l	94
82) 4-Chlorotoluene	11.451	91	237246	20.564	ug/l	99
83) tert-Butylbenzene	11.713	119	246305	20.341	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	253505	20.906	ug/l	99
85) sec-Butylbenzene	11.890	105	298610	20.104	ug/l	100
86) p-Isopropyltoluene	12.006	119	259007	20.503	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	137913	20.397	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	138291	19.508	ug/l	99
89) n-Butylbenzene	12.335	91	208446	19.547	ug/l	99
90) Hexachloroethane	12.536	117	38434	19.875	ug/l	88
91) 1,2-Dichlorobenzene	12.335	146	134977	19.960	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	18778	21.080	ug/l	88
93) 1,2,4-Trichlorobenzene	13.585	180	86590	20.371	ug/l	99
94) Hexachlorobutadiene	13.725	225	35381	18.966	ug/l	99
95) Naphthalene	13.774	128	277579	21.212	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	85675	20.022	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
Data File : VX035382.D
Acq On : 28 Apr 2023 11:44
Operator : JC/MD
Sample : VX0428WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01

Quant Time: Apr 28 15:40:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
Data File : VX035382.D
Acq On : 28 Apr 2023 11:44
Operator : JC/MD
Sample : VX0428WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01

Quant Time: Apr 28 15:40:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:168 Resp: 288171
Ion Ratio Lower Upper
168 100
99 57.5 45.7 68.5



#2
Dichlorodifluoromethane
Concen: 20.210 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 85 Resp: 61623
Ion Ratio Lower Upper
85 100
87 32.9 16.4 49.1









#3
Chloromethane
Concen: 17.589 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01

Tgt Ion: 50 Resp: 48789
Ion Ratio Lower Upper
50 100
52 32.5 25.9 38.9






#4
Vinyl Chloride
Concen: 18.515 ug/l
RT: 1.373 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 62 Resp: 51510
Ion Ratio Lower Upper
62 100
64 33.9 26.4 39.6



Abundance Scan 87 (1.617 min): VX035275.D\data.ms (-81)



#5

Bromomethane

Concen: 18.093 ug/l

RT: 1.617 min Scan# 81

Delta R.T. 0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :


VX0428WBS01

Tgt Ion: 94 Resp: 33456




Ion Ratio Lower Upper

94 100

96 98.1 77.0 115.6



Abundance Scan 99 (1.690 min): VX035275.D\data.ms (-93)



#6

Chloroethane

Concen: 18.330 ug/l

RT: 1.691 min Scan# 99

Delta R.T. 0.001 min

Lab File: VX035382.D


Acq: 28 Apr 2023 11:44


Tgt Ion: 64 Resp: 29825

Ion Ratio Lower Upper

64 100

66 34.8 26.2 39.4









#7
Trichlorofluoromethane
Concen: 20.162 ug/l
RT: 1.886 min Scan# 11
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:101 Resp: 87719
Ion Ratio Lower Upper
101 100
103 62.1 52.0 78.0



#8
Diethyl Ether
Concen: 18.931 ug/l
RT: 2.129 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 74 Resp: 30030
Ion Ratio Lower Upper
74 100
45 91.4 49.8 149.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.478 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:101 Resp: 56599
Ion Ratio Lower Upper
101 100
85 42.7 35.4 53.2
151 81.6 60.2 90.4



#10
Methyl Iodide
Concen: 16.020 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:142 Resp: 58752
Ion Ratio Lower Upper
142 100
127 43.7 34.0 51.0
141 14.1 11.1 16.7









#11
Tert butyl alcohol
Concen: 97.332 ug/l
RT: 3.013 min Scan# 31
Delta R.T. -0.012 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 59 Resp: 61613
Ion Ratio Lower Upper
59 100
57 0.0 8.1 12.1#



#12
1,1-Dichloroethene
Concen: 18.965 ug/l
RT: 2.318 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 96 Resp: 53080
Ion Ratio Lower Upper
96 100
61 164.8 128.8 193.2
98 66.7 51.4 77.0









#13
Acrolein
Concen: 90.954 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 56 Resp: 56622
Ion Ratio Lower Upper
56 100
55 70.6 56.6 85.0



#14
Allyl chloride
Concen: 18.655 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44


Tgt Ion: 41 Resp: 91884
Ion Ratio Lower Upper
41 100
39 63.5 52.4 78.6
76 34.4 26.0 39.0








#15
Acrylonitrile
Concen: 93.249 ug/l
RT: 3.068 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01



Tgt Ion: 53 Resp: 147637




Ion	Ratio	Lower	Upper
53	100		
52	83.1	66.2	99.2
51	36.3	28.3	42.5




#16
Acetone
Concen: 99.624 ug/l
RT: 2.398 min Scan# 215
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 43 Resp: 150428

Ion	Ratio	Lower	Upper
43	100		
58	31.8	24.7	37.1









#17
Carbon Disulfide
Concen: 18.430 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 76 Resp: 123655
Ion Ratio Lower Upper
76 100
78 9.4 7.2 10.8



#18
Methyl Acetate
Concen: 19.132 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 43 Resp: 111712
Ion Ratio Lower Upper
43 100
74 25.0 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 20.204 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 73 Resp: 194002
Ion Ratio Lower Upper
73 100
57 22.2 18.5 27.7



#20
Methylene Chloride
Concen: 17.817 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 84 Resp: 61685
Ion Ratio Lower Upper
84 100
49 120.6 103.8 155.8
51 37.6 31.8 47.8
86 65.7 51.4 77.2









#21
trans-1,2-Dichloroethene
Concen: 19.378 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 96 Resp: 58906
Ion Ratio Lower Upper
96 100
61 129.9 113.0 169.6
98 63.5 52.1 78.1



#22
Diisopropyl ether
Concen: 19.686 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 45 Resp: 191873
Ion Ratio Lower Upper
45 100
43 81.9 65.4 98.2
87 28.7 20.6 31.0
59 11.7 9.0 13.6









#23
 Vinyl Acetate
 Concen: 99.102 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. 0.001 min
 Lab File: VX035382.D
 Acq: 28 Apr 2023 11:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0428WBS01


Tgt Ion: 43 Resp: 663951
 Ion Ratio Lower Upper
 43 100
 86 11.2 8.5 12.7



#24
 1,1-Dichloroethane
 Concen: 19.634 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.000 min
 Lab File: VX035382.D
 Acq: 28 Apr 2023 11:44

Tgt Ion: 63 Resp: 107793
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.8 11.2
 100 5.1 2.4 7.1





#25

2-Butanone

Concen: 95.614 ug/l

RT: 4.568 min Scan# 571

Delta R.T. 0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :





VX0428WBS01

Tgt Ion: 43 Resp: 210726

Ion Ratio Lower Upper

43 100

72 26.6 19.3 28.9



#26

2,2-Dichloropropane

Concen: 20.167 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.001 min

Lab File: VX035382.D




Acq: 28 Apr 2023 11:44


Tgt Ion: 77 Resp: 86743

Ion Ratio Lower Upper

77 100

97 23.6 11.4 34.1





#27

cis-1,2-Dichloroethene

Concen: 19.447 ug/l

RT: 4.489 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :

VX0428WBS01





Tgt Ion: 96 Resp: 69487

Ion Ratio Lower Upper

96 100

61 141.3 0.0 300.4

98 65.2 0.0 129.0



#28

Bromochloromethane

Concen: 20.565 ug/l

RT: 4.897 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44




Tgt Ion: 49 Resp: 48115


Ion Ratio Lower Upper

49 100

129 2.1 0.0 4.0

130 88.5 58.2 87.2#





#29

Tetrahydrofuran

Concen: 91.966 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :

VX0428WBS01





Tgt Ion: 42 Resp: 129931

Ion Ratio Lower Upper

42 100

72 48.1 34.5 51.7

71 44.0 32.1 48.1



#30

Chloroform

Concen: 20.353 ug/l

RT: 5.092 min Scan# 657

Delta R.T. 0.000 min

Lab File: VX035382.D




Acq: 28 Apr 2023 11:44


Tgt Ion: 83 Resp: 112803

Ion Ratio Lower Upper

83 100

85 64.3 50.6 76.0









#31
Cyclohexane
Concen: 19.275 ug/l
RT: 5.470 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 56 Resp: 95058
Ion Ratio Lower Upper
56 100
69 32.1 23.7 35.5
84 90.6 68.5 102.7



#32
1,1,1-Trichloroethane
Concen: 20.264 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 97 Resp: 98424
Ion Ratio Lower Upper
97 100
99 64.3 51.1 76.7
61 43.0 36.5 54.7









#33
1,2-Dichloroethane-d4
Concen: 52.768 ug/l
RT: 5.952 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 65 Resp: 183455
Ion Ratio Lower Upper
65 100
67 51.5 0.0 104.0



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:114 Resp: 446481
Ion Ratio Lower Upper
114 100
63 19.9 0.0 42.2
88 16.2 0.0 33.2









#35
Dibromofluoromethane
Concen: 54.579 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:113 Resp: 149568
Ion Ratio Lower Upper
113 100
111 103.2 83.7 125.5
192 21.0 15.4 23.2



#36
1,1-Dichloropropene
Concen: 19.509 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.006 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 75 Resp: 83476
Ion Ratio Lower Upper
75 100
110 36.5 17.5 52.5
77 31.7 24.3 36.5









#37
Ethyl Acetate
Concen: 19.386 ug/l
RT: 4.714 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 43 Resp: 75762
Ion Ratio Lower Upper
43 100
61 15.1 10.2 15.4
70 12.4 8.2 12.4#



#38
Carbon Tetrachloride
Concen: 20.805 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 117 Resp: 84512
Ion Ratio Lower Upper
117 100
119 96.0 76.6 114.8
121 32.0 23.8 35.6





#39

Methylcyclohexane

Concen: 19.403 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. 0.001 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :

VX0428WBS01





Tgt Ion: 83 Resp: 101063

Ion Ratio Lower Upper

83 100

55 74.1 60.5 90.7

98 48.2 40.4 60.6



#40

Benzene

Concen: 19.855 ug/l

RT: 6.037 min Scan# 812

Delta R.T. 0.000 min

Lab File: VX035382.D




Acq: 28 Apr 2023 11:44


Tgt Ion: 78 Resp: 243711

Ion Ratio Lower Upper

78 100


77 22.9 19.0 28.4








#41
Methacrylonitrile
Concen: 19.375 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01



Tgt Ion: 41 Resp: 43388




Ion	Ratio	Lower	Upper
41	100		
39	56.8	45.0	67.6
67	75.9	55.6	83.4
52	30.5	23.2	34.8




#42
1,2-Dichloroethane
Concen: 20.300 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 62 Resp: 88843

Ion	Ratio	Lower	Upper
62	100		
98	10.3	0.0	18.0









#43
Isopropyl Acetate
Concen: 19.457 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 43 Resp: 127770
Ion Ratio Lower Upper
43 100
61 21.7 15.2 22.8
87 14.4 9.8 14.8



#44
Trichloroethene
Concen: 20.193 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.007 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:130 Resp: 68282
Ion Ratio Lower Upper
130 100
95 89.6 0.0 198.4









#45
1,2-Dichloropropane
Concen: 20.198 ug/l
RT: 7.427 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 63 Resp: 62505
Ion Ratio Lower Upper
63 100
65 31.6 24.6 37.0



#46
Dibromomethane
Concen: 20.328 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 93 Resp: 43200
Ion Ratio Lower Upper
93 100
95 81.6 66.9 100.3
174 106.9 75.4 113.0









#47
Bromodichloromethane
Concen: 20.641 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 83 Resp: 80820
Ion Ratio Lower Upper
83 100
85 61.4 52.2 78.4
127 8.7 6.6 9.8



#48
Methyl methacrylate
Concen: 19.904 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.007 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 41 Resp: 64115
Ion Ratio Lower Upper
41 100
69 86.2 62.2 93.4
39 56.6 47.7 71.5





#49

1,4-Dioxane

Concen: 409.510 ug/l

RT: 7.720 min Scan# 1088

Delta R.T. -0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :

VX0428WBS01





Tgt Ion: 88 Resp: 32154

Ion Ratio Lower Upper

88 100

43 29.0 27.7 41.5

58 66.8 56.8 85.2



#50

Toluene-d8

Concen: 53.670 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.001 min

Lab File: VX035382.D




Acq: 28 Apr 2023 11:44


Tgt Ion: 98 Resp: 554818

Ion Ratio Lower Upper

98 100

100 66.8 52.5 78.7









#51
4-Methyl-2-Pentanone
Concen: 98.756 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 43 Resp: 398427
Ion Ratio Lower Upper
43 100
58 41.1 30.6 46.0



#52
Toluene
Concen: 20.345 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 92 Resp: 158294
Ion Ratio Lower Upper
92 100
91 166.8 136.8 205.2





#53

t-1,3-Dichloropropene

Concen: 20.788 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :





VX0428WBS01

Tgt Ion: 75 Resp: 85001

Ion Ratio Lower Upper

75 100

77 30.7 25.4 38.2



#54

cis-1,3-Dichloropropene

Concen: 20.882 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44




Tgt Ion: 75 Resp: 96512


Ion Ratio Lower Upper

75 100

77 32.1 25.0 37.4

39 47.0 44.6 66.8









#55
1,1,2-Trichloroethane
Concen: 20.304 ug/l
RT: 9.153 min Scan# 11
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01

Tgt Ion: 97 Resp: 62979




Ion	Ratio	Lower	Upper
97	100		
83	81.3	69.7	104.5
85	52.9	43.7	65.5
99	62.1	49.7	74.5




#56
Ethyl methacrylate
Concen: 20.569 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 69 Resp: 95039

Ion	Ratio	Lower	Upper
69	100		
41	64.0	57.4	86.0
39	34.5	33.7	50.5





#57

1,3-Dichloropropane

Concen: 19.670 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :





VX0428WBS01

Tgt Ion: 76 Resp: 104355

Ion Ratio Lower Upper

76 100

78 32.9 25.8 38.8



#58

2-Chloroethyl Vinyl ether

Concen: 92.844 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX035382.D




Acq: 28 Apr 2023 11:44


Tgt Ion: 63 Resp: 214504

Ion Ratio Lower Upper

63 100

106 30.0 22.2 33.2





#59

2-Hexanone

Concen: 100.525 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.006 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :





VX0428WBS01

Tgt Ion: 43 Resp: 304677

Ion Ratio Lower Upper

43 100

58 56.1 26.3 78.9



#60

Dibromochloromethane

Concen: 21.022 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX035382.D




Acq: 28 Apr 2023 11:44


Tgt Ion: 129 Resp: 62473

Ion Ratio Lower Upper

129 100

127 76.7 39.0 117.0





#61

1,2-Dibromoethane

Concen: 20.648 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. -0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :





VX0428WBS01

Tgt Ion:107 Resp: 66023

Ion Ratio Lower Upper

107 100

109 93.2 75.3 112.9



#62

4-Bromofluorobenzene

Concen: 54.570 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44




Tgt Ion: 95 Resp: 212213


Ion Ratio Lower Upper

95 100

174 78.3 0.0 147.4

176 76.3 0.0 140.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:117 Resp: 395148
Ion Ratio Lower Upper
117 100
82 54.9 45.0 67.4
119 32.8 26.0 39.0



#64
Tetrachloroethene
Concen: 19.990 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:164 Resp: 60482
Ion Ratio Lower Upper
164 100
166 127.2 102.7 154.1
129 91.0 75.6 113.4
131 91.0 73.7 110.5









#65
Chlorobenzene
Concen: 19.700 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:112 Resp: 168417
Ion Ratio Lower Upper
112 100
114 32.1 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 20.708 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:131 Resp: 61613
Ion Ratio Lower Upper
131 100
133 94.9 47.1 141.4
119 61.7 32.6 98.0





#67

Ethyl Benzene

Concen: 20.313 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX035382.D

Acq: 28 Apr 2023 11:44

Instrument :

MSVOA_X

ClientSampleId :





VX0428WBS01

Tgt Ion: 91 Resp: 300745

Ion Ratio Lower Upper

91 100

106 30.3 25.6 38.4



#68

m/p-Xylenes

Concen: 40.538 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX035382.D




Acq: 28 Apr 2023 11:44


Tgt Ion: 106 Resp: 229975

Ion Ratio Lower Upper

106 100

91 205.9 164.0 246.0









#69
o-Xylene
Concen: 20.466 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:106 Resp: 113853
Ion Ratio Lower Upper
106 100
91 214.7 108.5 325.3



#70
Styrene
Concen: 20.910 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:104 Resp: 188601
Ion Ratio Lower Upper
104 100
78 51.8 43.4 65.2
103 53.8 44.2 66.2









#71
Bromoform
Concen: 20.905 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:173 Resp: 43263
Ion Ratio Lower Upper
173 100
175 46.9 24.1 72.4
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:152 Resp: 195798
Ion Ratio Lower Upper
152 100
115 70.5 45.0 135.0
150 163.1 0.0 352.4









#73
Isopropylbenzene
Concen: 21.039 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:105 Resp: 301130
Ion Ratio Lower Upper
105 100
120 26.6 13.3 39.9



#74
N-ethyl acetate
Concen: 20.489 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 43 Resp: 110207
Ion Ratio Lower Upper
43 100
70 44.5 32.1 48.1
55 25.0 18.3 27.5
61 26.0 18.8 28.2









#75
1,1,2,2-Tetrachloroethane
Concen: 20.766 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 83 Resp: 92250
Ion Ratio Lower Upper
83 100
131 10.8 4.8 14.3
85 63.4 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 24.057 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 75 Resp: 93959
Ion Ratio Lower Upper
75 100
77 33.8 21.6 64.8









#77
Bromobenzene
Concen: 20.707 ug/l
RT: 11.195 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:156 Resp: 73671
Ion Ratio Lower Upper
156 100
77 155.9 86.7 259.9
158 97.6 48.0 144.0



#78
n-propylbenzene
Concen: 20.373 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 91 Resp: 334888
Ion Ratio Lower Upper
91 100
120 23.7 11.5 34.5









#79
2-Chlorotoluene
Concen: 20.074 ug/l
RT: 11.366 min Scan# 1
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 91 Resp: 206199
Ion Ratio Lower Upper
91 100
126 34.4 16.3 48.9



#80
1,3,5-Trimethylbenzene
Concen: 21.121 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 105 Resp: 254895
Ion Ratio Lower Upper
105 100
120 48.0 23.7 71.1








#81
trans-1,4-Dichloro-2-butene
Concen: 18.602 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01

Tgt Ion: 75 Resp: 22433




Ion	Ratio	Lower	Upper
75	100		
53	112.3	82.6	124.0
89	46.7	36.9	55.3




#82
4-Chlorotoluene
Concen: 20.564 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 91 Resp: 237246

Ion	Ratio	Lower	Upper
91	100		
126	29.1	14.2	42.6









#83
tert-Butylbenzene
Concen: 20.341 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:119 Resp: 246305
Ion Ratio Lower Upper
119 100
91 59.3 30.6 92.0
134 23.9 11.8 35.4



#84
1,2,4-Trimethylbenzene
Concen: 20.906 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:105 Resp: 253505
Ion Ratio Lower Upper
105 100
120 44.0 22.4 67.0









#85
 sec-Butylbenzene
 Concen: 20.104 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX035382.D
 Acq: 28 Apr 2023 11:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0428WBS01


Tgt Ion:105 Resp: 298610
 Ion Ratio Lower Upper
 105 100
 134 20.1 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 20.503 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX035382.D
 Acq: 28 Apr 2023 11:44

Tgt Ion:119 Resp: 259007
 Ion Ratio Lower Upper
 119 100
 134 26.0 13.2 39.5
 91 24.0 12.9 38.7









#87
1,3-Dichlorobenzene
Concen: 20.397 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:146 Resp: 137913
Ion Ratio Lower Upper
146 100
111 40.8 20.9 62.8
148 64.1 31.6 95.0



#88
1,4-Dichlorobenzene
Concen: 19.508 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:146 Resp: 138291
Ion Ratio Lower Upper
146 100
111 41.0 20.6 61.9
148 64.1 31.8 95.4









#89
n-Butylbenzene
Concen: 19.547 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion: 91 Resp: 208446
Ion Ratio Lower Upper
91 100
92 54.9 27.7 83.0
134 26.9 13.0 38.9



#90
Hexachloroethane
Concen: 19.875 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 117 Resp: 38434
Ion Ratio Lower Upper
117 100
201 78.1 34.4 103.1









#91
1,2-Dichlorobenzene
Concen: 19.960 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:146 Resp: 134977
Ion Ratio Lower Upper
146 100
111 42.1 22.0 66.0
148 64.1 32.0 96.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.080 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.001 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion: 75 Resp: 18778
Ion Ratio Lower Upper
75 100
155 86.9 38.4 115.1
157 112.2 49.9 149.7









#93
1,2,4-Trichlorobenzene
Concen: 20.371 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44




Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01


Tgt Ion:180 Resp: 86590
Ion Ratio Lower Upper
180 100
182 95.4 47.0 141.0
145 31.8 15.9 47.7



#94
Hexachlorobutadiene
Concen: 18.966 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:225 Resp: 35381
Ion Ratio Lower Upper
225 100
223 64.6 32.1 96.5
227 65.4 31.8 95.4









#95
Naphthalene
Concen: 21.212 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBS01

Tgt Ion:	128	Resp:	277579
Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.2	15.4
129	11.1	8.6	13.0



#96
1,2,3-Trichlorobenzene
Concen: 20.022 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX035382.D
Acq: 28 Apr 2023 11:44

Tgt Ion:	180	Resp:	85675
Ion	Ratio	Lower	Upper
180	100		
182	96.4	48.0	144.2
145	34.0	17.7	53.1

