

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
 Data File : VX035389.D
 Acq On : 28 Apr 2023 14:33
 Operator : JC/MD
 Sample : 02386-02MEDL 40X
 Misc : 3.32g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPA-1WC-23-04MEDL

Quant Time: Apr 29 00:37:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	229234	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	377161	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	347462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	137201	49.610	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.220%
35) Dibromofluoromethane	5.385	113	110159	47.587	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	8.647	98	445994	51.072	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.140%
62) 4-Bromofluorobenzene	11.079	95	174670	53.171	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.340%
Target Compounds						
					Qvalue	
65) Chlorobenzene	10.079	112	28717	3.820	ug/l	98
67) Ethyl Benzene	10.195	91	131462	10.098	ug/l	98
68) m/p-Xylenes	10.299	106	109223	21.895	ug/l	100
69) o-Xylene	10.640	106	31062	6.350	ug/l	98
73) Isopropylbenzene	10.963	105	18521	1.465	ug/l	98
78) n-propylbenzene	11.305	91	36326	2.502	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	49743	4.667	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	159076	14.854	ug/l	99
86) p-Isopropyltoluene	12.006	119	19904	1.784	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	8730	1.462	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	44848	7.163	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	47871	8.015	ug/l	100
93) 1,2,4-Trichlorobenzene	13.591	180	14906	3.971	ug/l	93
95) Naphthalene	13.774	128	67766	5.864	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	4046	1.071	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
Data File : VX035389.D
Acq On : 28 Apr 2023 14:33
Operator : JC/MD
Sample : 02386-02MEDL 40X
Misc : 3.32g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPA-1WC-23-04MEDL

Quant Time: Apr 29 00:37:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

