

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
 Data File : VX035403.D
 Acq On : 28 Apr 2023 19:58
 Operator : JC/MD
 Sample : 02522-07DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6DL

Quant Time: Apr 29 00:43:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	249852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	409955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	183641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	145979	48.429	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.860%
35) Dibromofluoromethane	5.385	113	123734	49.175	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.360%
50) Toluene-d8	8.647	98	481059	50.681	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.360%
62) 4-Bromofluorobenzene	11.079	95	187338	52.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.940%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.470	56	6232	1.457	ug/l #	92
39) Methylcyclohexane	7.379	83	12419	2.597	ug/l	91
40) Benzene	6.037	78	132902	11.792	ug/l	98
52) Toluene	8.720	92	20251	2.835	ug/l	99
67) Ethyl Benzene	10.195	91	79082	5.646	ug/l	98
68) m/p-Xylenes	10.299	106	351381	65.473	ug/l	100
69) o-Xylene	10.640	106	153874	29.239	ug/l	98
73) Isopropylbenzene	10.963	105	4943	0.368	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	73482	6.492	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	242763	21.346	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
Data File : VX035403.D
Acq On : 28 Apr 2023 19:58
Operator : JC/MD
Sample : 02522-07DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6DL

Quant Time: Apr 29 00:43:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

