

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042823\  
 Data File : VX035404.D  
 Acq On : 28 Apr 2023 20:21  
 Operator : JC/MD  
 Sample : 02495-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PZ-02

Quant Time: Apr 29 00:44:27 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X042523W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 25 16:31:40 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 238419   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 394894   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 363087   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 173207   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 142132   | 49.413   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 98.820%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 120877   | 49.872   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 99.740%  |
| 50) Toluene-d8              | 8.647  | 98    | 463850   | 50.732   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 101.460% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 177388   | 51.574   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 103.140% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 2.392  | 43    | 6912     | 5.533    | ug/l # | 84       |
| 22) Diisopropyl ether       | 3.757  | 45    | 1659     | 0.206    | ug/l # | 61       |
| 68) m/p-Xylenes             | 10.305 | 106   | 1333     | 0.256    | ug/l   | 89       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042823\  
Data File : VX035404.D  
Acq On : 28 Apr 2023 20:21  
Operator : JC/MD  
Sample : 02495-03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 27 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PZ-02

Quant Time: Apr 29 00:44:27 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X042523W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 25 16:31:40 2023  
Response via : Initial Calibration

