

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
 Data File : VX035410.D
 Acq On : 28 Apr 2023 23:26
 Operator : JC/MD
 Sample : VX0428WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0428WBL02

Quant Time: Apr 29 02:02:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	262544	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	432281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	395435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	188780	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	150811	47.613	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.220%
35) Dibromofluoromethane	5.385	113	128799	48.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.080%
50) Toluene-d8	8.647	98	502859	50.242	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.480%
62) 4-Bromofluorobenzene	11.079	95	193724	51.452	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
Data File : VX035410.D
Acq On : 28 Apr 2023 23:26
Operator : JC/MD
Sample : VX0428WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0428WBL02

Quant Time: Apr 29 02:02:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

