

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
 Data File : VX035413.D
 Acq On : 29 Apr 2023 00:36
 Operator : JC/MD
 Sample : PB152434TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB152434TB

Quant Time: Apr 29 02:04:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	251972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	414442	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	182711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148753	48.934	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.860%
35) Dibromofluoromethane	5.385	113	128197	50.397	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	8.647	98	486873	50.738	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.480%
62) 4-Bromofluorobenzene	11.079	95	186910	51.779	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.560%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	23278	17.631	ug/l	98
20) Methylene Chloride	2.788	84	9193	3.037	ug/l	94
25) 2-Butanone	4.580	43	16998	8.821	ug/l #	88
37) Ethyl Acetate	4.714	43	47831	13.185	ug/l #	96
43) Isopropyl Acetate	6.336	43	185802	30.481	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042823\
Data File : VX035413.D
Acq On : 29 Apr 2023 00:36
Operator : JC/MD
Sample : PB152434TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB152434TB

Quant Time: Apr 29 02:04:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

