

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042825\
 Data File : VX045954.D
 Acq On : 28 Apr 2025 14:17
 Operator : JC/MD
 Sample : Q1874-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT1491

Quant Time: Apr 29 01:34:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130914	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67804	56.143	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.280%
35) Dibromofluoromethane	5.379	113	48839	52.581	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.160%
50) Toluene-d8	8.647	98	168274	51.904	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.800%
62) 4-Bromofluorobenzene	11.079	95	64304	54.453	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.900%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	48205	97.204	ug/l	99
18) Methyl Acetate	2.709	43	1899	1.663	ug/l	99
20) Methylene Chloride	2.782	84	2462	2.652	ug/l	95
25) 2-Butanone	4.586	43	3787	5.217	ug/l	92
43) Isopropyl Acetate	6.348	43	11292	4.714	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042825\
Data File : VX045954.D
Acq On : 28 Apr 2025 14:17
Operator : JC/MD
Sample : Q1874-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT1491

Quant Time: Apr 29 01:34:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

