

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042825\
 Data File : VX045956.D
 Acq On : 28 Apr 2025 15:03
 Operator : JC/MD
 Sample : Q1875-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-25-0053

Quant Time: Apr 29 01:34:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125299	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63209	54.843	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.680%	
35) Dibromofluoromethane	5.379	113	45238	50.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.780%	
50) Toluene-d8	8.647	98	161618	52.085	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.180%	
62) 4-Bromofluorobenzene	11.079	95	60237	53.295	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.600%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.983	59	958	6.185	ug/l	98
16) Acetone	2.386	43	148993	314.818	ug/l	98
18) Methyl Acetate	2.703	43	2309	2.119	ug/l #	83
20) Methylene Chloride	2.782	84	2009	2.267	ug/l #	89
25) 2-Butanone	4.580	43	5009	7.230	ug/l	98
43) Isopropyl Acetate	6.342	43	5597	2.441	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042825\
Data File : VX045956.D
Acq On : 28 Apr 2025 15:03
Operator : JC/MD
Sample : Q1875-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-25-0053

Quant Time: Apr 29 01:34:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

