

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042825\
 Data File : VX045966.D
 Acq On : 28 Apr 2025 18:55
 Operator : JC/MD
 Sample : Q1900-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-2

Quant Time: Apr 29 01:36:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	64585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130657	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126880	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65372	55.349	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.700%
35) Dibromofluoromethane	5.379	113	47526	51.268	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.540%
50) Toluene-d8	8.646	98	166699	51.520	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.040%
62) 4-Bromofluorobenzene	11.079	95	63637	53.995	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.980%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	55692	114.831	ug/l	97
18) Methyl Acetate	2.702	43	1967	1.762	ug/l	93
20) Methylene Chloride	2.788	84	3235	3.563	ug/l	98
25) 2-Butanone	4.586	43	4592	6.468	ug/l #	87
43) Isopropyl Acetate	6.348	43	11008	4.605	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042825\
Data File : VX045966.D
Acq On : 28 Apr 2025 18:55
Operator : JC/MD
Sample : Q1900-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-2

Quant Time: Apr 29 01:36:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

