

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042919\
 Data File : VX009199.D
 Acq On : 29 Apr 2019 11:13
 Operator : JC/SP
 Sample : VX0429WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBL01

Quant Time: Apr 30 07:29:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	226718	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	366877	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	316432	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	136570	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	147101	47.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.06%	
35) Dibromofluoromethane	5.49	113	110357	47.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.70%	
50) Toluene-d8	8.71	98	448630	48.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.36%	
62) 4-Bromofluorobenzene	11.13	95	150910	44.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.20%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042919\
Data File : VX009199.D
Acq On : 29 Apr 2019 11:13
Operator : JC/SP
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01

Quant Time: Apr 30 07:29:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

