

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042919\
 Data File : VX009217.D
 Acq On : 29 Apr 2019 19:06
 Operator : JC/SP
 Sample : K2594-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021MW173-190423

Quant Time: Apr 30 08:22:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	206856	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	337553	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	293984	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	124311	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	136364	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
35) Dibromofluoromethane	5.50	113	102339	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
50) Toluene-d8	8.71	98	421316	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.13	95	138239	44.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.82%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3286	1.295	ug/l	97
16) Acetone	2.45	43	2162	1.426	ug/l #	87
27) cis-1,2-Dichloroethene	4.60	96	6513	2.592	ug/l	90
65) Chlorobenzene	10.13	112	1661	0.282	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042919\
Data File : VX009217.D
Acq On : 29 Apr 2019 19:06
Operator : JC/SP
Sample : K2594-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021MW173-190423

Quant Time: Apr 30 08:22:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

