

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042919\
 Data File : VX009224.D
 Acq On : 29 Apr 2019 21:49
 Operator : JC/SP
 Sample : K2594-13DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0514-190423DL

Quant Time: Apr 30 08:54:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	215135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	345263	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	304880	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	134116	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	141560	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
35) Dibromofluoromethane	5.50	113	104380	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
50) Toluene-d8	8.71	98	434694	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	11.13	95	146497	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	9454	3.618	ug/l	89
44) Trichloroethene	7.21	130	1728	0.720	ug/l	95
64) Tetrachloroethene	9.34	164	180408	86.264	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042919\
Data File : VX009224.D
Acq On : 29 Apr 2019 21:49
Operator : JC/SP
Sample : K2594-13DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0514-190423DL

Quant Time: Apr 30 08:54:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

