

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042919\
 Data File : VX009225.D
 Acq On : 29 Apr 2019 22:12
 Operator : JC/SP
 Sample : K2594-14 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0515-190423

Quant Time: Apr 30 08:55:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	211925	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	343184	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	302957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132859	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	139732	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
35) Dibromofluoromethane	5.50	113	103737	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
50) Toluene-d8	8.71	98	429627	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	11.13	95	145139	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	1405	0.546	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042919\
Data File : VX009225.D
Acq On : 29 Apr 2019 22:12
Operator : JC/SP
Sample : K2594-14 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0515-190423

Quant Time: Apr 30 08:55:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

