

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042919\
 Data File : VX009226.D
 Acq On : 29 Apr 2019 22:36
 Operator : JC/SP
 Sample : K2594-10 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021RW245-190423

Quant Time: Apr 30 09:07:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	195721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	315716	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	284043	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130675	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	128105	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
35) Dibromofluoromethane	5.50	113	94058	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
50) Toluene-d8	8.71	98	396034	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
62) 4-Bromofluorobenzene	11.13	95	142916	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
Target Compounds						
4) Vinyl Chloride	1.41	62	28912	12.041	ug/l	97
24) 1,1-Dichloroethane	3.69	63	2468	0.581	ug/l #	89
27) cis-1,2-Dichloroethene	4.59	96	156971	66.032	ug/l	90
65) Chlorobenzene	10.14	112	19289	3.395	ug/l	98
91) 1,2-Dichlorobenzene	12.39	146	1828	0.427	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042919\
Data File : VX009226.D
Acq On : 29 Apr 2019 22:36
Operator : JC/SP
Sample : K2594-10 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021RW245-190423

Quant Time: Apr 30 09:07:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

