

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042919\
 Data File : VX009227.D
 Acq On : 29 Apr 2019 22:59
 Operator : JC/SP
 Sample : K2594-12 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021RW246-190423

Quant Time: Apr 30 09:10:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	197142	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	313990	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	282711	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	123345	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	128302	47.67	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.34%	
35) Dibromofluoromethane	5.49	113	94257	47.75	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.50%	
50) Toluene-d8	8.71	98	399101	50.08	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.16%	
62) 4-Bromofluorobenzene	11.13	95	140057	48.37	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.74%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.41	62	789910	326.603	ug/l	98
12) 1,1-Dichloroethene	2.36	96	797	0.431	ug/l #	85
24) 1,1-Dichloroethane	3.70	63	2564	0.600	ug/l #	94
27) cis-1,2-Dichloroethene	4.59	96	57269	23.917	ug/l	91
39) Methylcyclohexane	7.46	83	803	0.224	ug/l #	83
40) Benzene	6.14	78	63523	7.034	ug/l	99
52) Toluene	8.78	92	2900	0.528	ug/l	95
64) Tetrachloroethene	9.34	164	1951	1.006	ug/l	95
65) Chlorobenzene	10.14	112	654517	115.731	ug/l	99
73) Isopropylbenzene	11.02	105	2291	0.255	ug/l	99
83) tert-Butylbenzene	11.77	119	7542	1.027	ug/l	88
85) sec-Butylbenzene	11.94	105	5261	0.604	ug/l #	57
87) 1,3-Dichlorobenzene	12.02	146	3580	0.902	ug/l	96
88) 1,4-Dichlorobenzene	12.10	146	12762	3.190	ug/l	90
91) 1,2-Dichlorobenzene	12.39	146	36731	9.087	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042919\
Data File : VX009227.D
Acq On : 29 Apr 2019 22:59
Operator : JC/SP
Sample : K2594-12 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021RW246-190423

Quant Time: Apr 30 09:10:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

