

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042919\
 Data File : VX009228.D
 Acq On : 29 Apr 2019 23:22
 Operator : JC/SP
 Sample : K2594-02 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021MW352-190423

Quant Time: May 01 10:28:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:03:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	214667	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	345057	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	303382	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132193	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	140099	47.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.62%	
35) Dibromofluoromethane	5.50	113	103112	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
50) Toluene-d8	8.71	98	429796	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
62) 4-Bromofluorobenzene	11.13	95	145171	45.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.24%	
Target Compounds						
12) 1,1-Dichloroethene	2.37	96	826	0.410	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	978	0.448	ug/l	81
27) cis-1,2-Dichloroethene	4.60	96	18262	7.004	ug/l	91
44) Trichloroethene	7.21	130	3361	1.401	ug/l	94
64) Tetrachloroethene	9.34	164	406575	195.369	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042919\
Data File : VX009228.D
Acq On : 29 Apr 2019 23:22
Operator : JC/SP
Sample : K2594-02 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021MW352-190423

Quant Time: May 01 10:28:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:03:50 2019
Response via : Initial Calibration

