


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX042920\
 Data File : VX015968.D
 Acq On : 29 Apr 2020 09:58
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042820W.M
 Title : SW846 8260
 Last Update : Tue Apr 28 17:30:11 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1639

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	10238	PASS
75	95	30	60	52.8	25954	PASS
95	95	100	100	100.0	49200	PASS
96	95	5	9	6.6	3259	PASS
173	174	0.00	2	1.1	467	PASS
174	95	50	100	84.3	41456	PASS
175	174	5	9	7.7	3201	PASS
176	174	95	101	95.8	39725	PASS
177	176	5	9	7.1	2823	PASS