

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028414.D
 Acq On : 29 Apr 2022 11:17
 Operator : JC/MD
 Sample : VX0429WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBL01

Quant Time: Apr 29 15:03:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X042822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 29 06:09:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	71867	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	396059	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	342881	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	142096	31.090	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.633%
60) 4-Bromofluorobenzene	11.079	95	152924	28.679	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.600%
63) Toluene-d8	8.647	98	452769	31.171	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.900%
Target Compounds						
18) Methylene Chloride	2.788	84	13435	3.859	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028414.D
Acq On : 29 Apr 2022 11:17
Operator : JC/MD
Sample : VX0429WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL01

Quant Time: Apr 29 15:03:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X042822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 29 06:09:36 2022
Response via : Initial Calibration

