

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028420.D
 Acq On : 29 Apr 2022 13:52
 Operator : JC/MD
 Sample : N2550-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PUMP-STATION-207ST-TRAIN-FACILITYRE

Quant Time: Apr 29 15:06:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X042822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 29 06:09:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	57254	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	319426	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	281473	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	122632	33.680	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	112.267%#
60) 4-Bromofluorobenzene	11.079	95	123296	28.167	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.900%
63) Toluene-d8	8.653	98	376509	31.576	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	105.267%
Target Compounds						
15) Acetone	2.380	58	1281	2.883	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028420.D
Acq On : 29 Apr 2022 13:52
Operator : JC/MD
Sample : N2550-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PUMP-STATION-207ST-TRAIN-FACILITYRE

Quant Time: Apr 29 15:06:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X042822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 29 06:09:36 2022
Response via : Initial Calibration

