

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028422.D
 Acq On : 29 Apr 2022 14:39
 Operator : JC/MD
 Sample : N2563-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORM-WATER-042622

Quant Time: Apr 30 08:16:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X042822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 29 06:09:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	53740	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	301869	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	269828	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	114290	33.441	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	111.467%#
60) 4-Bromofluorobenzene	11.079	95	117810	28.075	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.600%
63) Toluene-d8	8.653	98	353736	30.946	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.167%
Target Compounds						
					Qvalue	
15) Acetone	2.380	58	3161	7.580	ug/l	74
18) Methylene Chloride	2.788	84	1049	0.403	ug/l	95
34) Benzene	6.043	78	2782	0.272	ug/l #	69
62) Toluene	8.720	91	10972	0.938	ug/l	97
66) Ethyl Benzene	10.201	91	2921	0.212	ug/l	94
67) m/p-Xylenes	10.305	106	3141	0.568	ug/l	96
68) o-Xylene	10.646	106	1698	0.302	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028422.D
Acq On : 29 Apr 2022 14:39
Operator : JC/MD
Sample : N2563-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORM-WATER-042622

Quant Time: Apr 30 08:16:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X042822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 29 06:09:36 2022
Response via : Initial Calibration

