

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028426.D
 Acq On : 29 Apr 2022 16:13
 Operator : JC/MD
 Sample : N2640-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MWBR3D-B-480-500-042922

Quant Time: Apr 30 09:15:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	416139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	714282	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	687385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	324842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	292032	53.821	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.640%	
35) Dibromofluoromethane	5.391	113	268137	53.529	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.060%	
50) Toluene-d8	8.653	98	983433	51.116	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.240%	
62) 4-Bromofluorobenzene	11.079	95	356589	48.474	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.940%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	222795	69.664	ug/l	98
16) Acetone	2.380	43	11156	5.366	ug/l	95
24) 1,1-Dichloroethane	3.611	63	25766	3.570	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	230411	44.429	ug/l	84
42) 1,2-Dichloroethane	6.092	62	14279	2.011	ug/l	98
44) Trichloroethene	7.129	130	194005	34.978	ug/l	97
49) 1,4-Dioxane	7.659	88	4510	34.030	ug/l	93
52) Toluene	8.720	92	207272	15.598	ug/l	98
94) Hexachlorobutadiene	13.725	225	758	0.263	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028426.D
Acq On : 29 Apr 2022 16:13
Operator : JC/MD
Sample : N2640-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MWBR3D-B-480-500-042922

Quant Time: Apr 30 09:15:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

