

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028428.D
 Acq On : 29 Apr 2022 17:00
 Operator : JC/MD
 Sample : N2640-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-042922

Quant Time: Apr 30 09:16:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	394116	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	685536	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	657248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	304262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	282923	55.056	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.120%
35) Dibromofluoromethane	5.385	113	251108	52.231	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.460%
50) Toluene-d8	8.653	98	947101	51.291	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.580%
62) 4-Bromofluorobenzene	11.079	95	337697	47.831	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.660%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	14776	8.063	ug/l	94
20) Methylene Chloride	2.788	84	1369	0.359	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028428.D
Acq On : 29 Apr 2022 17:00
Operator : JC/MD
Sample : N2640-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-042922

Quant Time: Apr 30 09:16:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

