

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028430.D
 Acq On : 29 Apr 2022 17:46
 Operator : JC/MD
 Sample : N2640-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MWBR3D-B-480-500-042922-FD

Quant Time: Apr 30 09:16:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	391358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	673980	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	658473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	308407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	303280	59.433	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 118.860%		
35) Dibromofluoromethane	5.391	113	269102	56.934	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 113.860%		
50) Toluene-d8	8.653	98	997831	54.965	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 109.940%		
62) 4-Bromofluorobenzene	11.079	95	366915	52.860	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 105.720%		
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	197060	65.519	ug/l	96
16) Acetone	2.380	43	10619	5.448	ug/l	89
24) 1,1-Dichloroethane	3.611	63	23677	3.488	ug/l #	95
27) cis-1,2-Dichloroethene	4.495	96	210009	43.059	ug/l	84
30) Chloroform	5.099	83	3467	0.425	ug/l	97
42) 1,2-Dichloroethane	6.092	62	13813	2.062	ug/l	98
44) Trichloroethene	7.129	130	163679	31.275	ug/l	98
49) 1,4-Dioxane	7.665	88	4448	35.569	ug/l #	88
52) Toluene	8.720	92	178221	14.214	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028430.D
Acq On : 29 Apr 2022 17:46
Operator : JC/MD
Sample : N2640-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MWBR3D-B-480-500-042922-FD

Quant Time: Apr 30 09:16:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

