

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028437.D
 Acq On : 29 Apr 2022 20:28
 Operator : JC/MD
 Sample : N2609-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-4

Quant Time: Apr 30 09:18:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	392653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	687414	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	710369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	326478	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	289748	56.594	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.180%
35) Dibromofluoromethane	5.391	113	259956	53.924	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.840%
50) Toluene-d8	8.653	98	988835	53.405	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.820%
62) 4-Bromofluorobenzene	11.079	95	365837	51.675	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.340%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	13686	7.397	ug/l	95
20) Methylene Chloride	2.788	84	7145	1.881	ug/l	95
43) Isopropyl Acetate	6.342	43	25310	2.437	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028437.D
Acq On : 29 Apr 2022 20:28
Operator : JC/MD
Sample : N2609-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-4

Quant Time: Apr 30 09:18:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

