

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028442.D
 Acq On : 29 Apr 2022 23:11
 Operator : JC/MD
 Sample : VX0429WBL03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0429WBL03

Quant Time: Apr 30 09:20:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	392796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	677906	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	652276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	306290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	289931	56.609	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 113.220%	
35) Dibromofluoromethane	5.385	113	259662	54.619	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.240%	
50) Toluene-d8	8.653	98	964719	52.834	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.660%	
62) 4-Bromofluorobenzene	11.085	95	346860	49.681	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028442.D
Acq On : 29 Apr 2022 23:11
Operator : JC/MD
Sample : VX0429WBL03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0429WBL03

Quant Time: Apr 30 09:20:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

