

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028446.D
 Acq On : 30 Apr 2022 00:44
 Operator : JC/MD
 Sample : PB144452ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144452ZHE#01

Quant Time: Apr 30 09:21:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	392375	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	690316	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	710814	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	318308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	284233	55.556	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 111.120%	
35) Dibromofluoromethane	5.391	113	252399	52.136	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.280%	
50) Toluene-d8	8.653	98	979542	52.681	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.360%	
62) 4-Bromofluorobenzene	11.079	95	356421	50.133	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.260%	
Target Compounds						Qvalue
16) Acetone	2.380	43	13249	7.122	ug/l	97
20) Methylene Chloride	2.794	84	19986	5.265	ug/l	96
43) Isopropyl Acetate	6.348	43	27138	2.602	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028446.D
Acq On : 30 Apr 2022 00:44
Operator : JC/MD
Sample : PB144452ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144452ZHE#01

Quant Time: Apr 30 09:21:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

