

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028448.D
 Acq On : 30 Apr 2022 01:30
 Operator : JC/MD
 Sample : PB144452ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144452ZHE#03

Quant Time: Apr 30 09:21:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	384667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	669005	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	682558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	310844	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	286649	57.151	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	114.300%
35) Dibromofluoromethane	5.391	113	257433	54.870	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.740%
50) Toluene-d8	8.653	98	989597	54.917	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.840%
62) 4-Bromofluorobenzene	11.079	95	356856	51.793	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.580%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	13036	7.153	ug/l	97
20) Methylene Chloride	2.788	84	20287	5.451	ug/l	94
43) Isopropyl Acetate	6.342	43	28498	2.820	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028448.D
Acq On : 30 Apr 2022 01:30
Operator : JC/MD
Sample : PB144452ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144452ZHE#03

Quant Time: Apr 30 09:21:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

