

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028450.D
 Acq On : 30 Apr 2022 02:17
 Operator : JC/MD
 Sample : PB144452ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144452ZHE#05

Quant Time: Apr 30 09:22:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	388419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	694773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	717815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	333018	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	299574	59.151	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 118.300%	
35) Dibromofluoromethane	5.391	113	264864	54.360	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 108.720%	
50) Toluene-d8	8.653	98	1027721	54.918	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 109.840%	
62) 4-Bromofluorobenzene	11.079	95	380940	53.238	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 106.480%	
Target Compounds						Qvalue
16) Acetone	2.380	43	10055	5.133	ug/l	98
20) Methylene Chloride	2.788	84	9246	2.460	ug/l	91
43) Isopropyl Acetate	6.342	43	29403	2.801	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028450.D
Acq On : 30 Apr 2022 02:17
Operator : JC/MD
Sample : PB144452ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144452ZHE#05

Quant Time: Apr 30 09:22:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

