

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028452.D
 Acq On : 30 Apr 2022 03:03
 Operator : JC/MD
 Sample : PB144452ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144452ZHE#07

Quant Time: Apr 30 09:22:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	375335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	674949	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	694016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	311862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	289327	59.120	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	118.240%
35) Dibromofluoromethane	5.391	113	256141	54.114	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.220%
50) Toluene-d8	8.653	98	987721	54.330	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.660%
62) 4-Bromofluorobenzene	11.079	95	359710	51.748	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.500%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	9697	5.120	ug/l	95
20) Methylene Chloride	2.794	84	9574	2.637	ug/l	93
43) Isopropyl Acetate	6.348	43	28253	2.771	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028452.D
Acq On : 30 Apr 2022 03:03
Operator : JC/MD
Sample : PB144452ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144452ZHE#07

Quant Time: Apr 30 09:22:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

