

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028453.D
 Acq On : 30 Apr 2022 03:27
 Operator : JC/MD
 Sample : PB144452ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144452ZHE#08

Quant Time: Apr 30 09:22:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	379102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	674422	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	690157	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	314259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	288156	58.295	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.600%
35) Dibromofluoromethane	5.391	113	258879	54.735	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.480%
50) Toluene-d8	8.653	98	991038	54.555	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.120%
62) 4-Bromofluorobenzene	11.079	95	362841	52.239	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.480%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	10363	5.499	ug/l	95
20) Methylene Chloride	2.788	84	9623	2.624	ug/l	98
43) Isopropyl Acetate	6.342	43	28713	2.818	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028453.D
Acq On : 30 Apr 2022 03:27
Operator : JC/MD
Sample : PB144452ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144452ZHE#08

Quant Time: Apr 30 09:22:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

