

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028454.D
 Acq On : 30 Apr 2022 03:50
 Operator : JC/MD
 Sample : PB144452ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144452ZHE#09

Quant Time: Apr 30 09:23:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	386817	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	690469	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	720514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	338331	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	293908	58.273	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 116.540%			
35) Dibromofluoromethane	5.391	113	260996	53.900	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 107.800%			
50) Toluene-d8	8.653	98	1009832	54.298	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.600%			
62) 4-Bromofluorobenzene	11.079	95	374445	52.657	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 105.320%			
Target Compounds						
16) Acetone	2.380	43	10498	5.449	ug/l	88
20) Methylene Chloride	2.794	84	9435	2.521	ug/l	93
43) Isopropyl Acetate	6.342	43	28231	2.707	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028454.D
Acq On : 30 Apr 2022 03:50
Operator : JC/MD
Sample : PB144452ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144452ZHE#09

Quant Time: Apr 30 09:23:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

