

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028455.D
 Acq On : 30 Apr 2022 04:13
 Operator : JC/MD
 Sample : PB144452ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144452ZHE#10

Quant Time: Apr 30 09:23:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	376427	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	665644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	689511	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	304055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	287120	58.498	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 117.000%			
35) Dibromofluoromethane	5.391	113	254312	54.479	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 108.960%			
50) Toluene-d8	8.653	98	987380	55.071	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 110.140%			
62) 4-Bromofluorobenzene	11.079	95	356906	52.062	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 104.120%			
Target Compounds						Qvalue
20) Methylene Chloride	2.788	84	6842	1.879	ug/l	98
43) Isopropyl Acetate	6.342	43	28022	2.787	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028455.D
Acq On : 30 Apr 2022 04:13
Operator : JC/MD
Sample : PB144452ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144452ZHE#10

Quant Time: Apr 30 09:23:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

