

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028457.D
 Acq On : 30 Apr 2022 04:59
 Operator : JC/MD
 Sample : PB144452ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144452ZHE#12

Quant Time: Apr 30 09:23:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	393425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	692261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	718905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	325984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	292677	57.054	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	114.100%
35) Dibromofluoromethane	5.391	113	261217	53.806	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.620%
50) Toluene-d8	8.653	98	1013789	54.370	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.740%
62) 4-Bromofluorobenzene	11.079	95	372338	52.225	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.440%
Target Compounds						Qvalue
20) Methylene Chloride	2.794	84	6426	1.688	ug/l	98
43) Isopropyl Acetate	6.348	43	27540	2.633	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028457.D
Acq On : 30 Apr 2022 04:59
Operator : JC/MD
Sample : PB144452ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144452ZHE#12

Quant Time: Apr 30 09:23:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

