

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028460.D
 Acq On : 30 Apr 2022 06:09
 Operator : JC/MD
 Sample : N2588-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-2

Quant Time: Apr 30 09:24:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	382699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	670005	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	700808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	327386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	292620	58.642	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.280%
35) Dibromofluoromethane	5.385	113	258898	55.100	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	110.200%
50) Toluene-d8	8.653	98	993724	55.064	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.120%
62) 4-Bromofluorobenzene	11.079	95	371103	53.781	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.560%
Target Compounds						Qvalue
16) Acetone	2.380	43	14833	8.383	ug/l	93
20) Methylene Chloride	2.794	84	21821	5.894	ug/l	97
43) Isopropyl Acetate	6.342	43	25382	2.508	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028460.D
Acq On : 30 Apr 2022 06:09
Operator : JC/MD
Sample : N2588-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-2

Quant Time: Apr 30 09:24:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

