

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028464.D
 Acq On : 30 Apr 2022 07:42
 Operator : JC/MD
 Sample : N2616-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-18D

Quant Time: Apr 30 09:25:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	381364	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	679272	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	697510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	317919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	290099	58.340	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 116.680%	
35) Dibromofluoromethane	5.391	113	255904	53.720	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.440%	
50) Toluene-d8	8.653	98	975874	53.337	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.680%	
62) 4-Bromofluorobenzene	11.079	95	359687	51.415	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.820%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	11013	5.888	ug/l	92
20) Methylene Chloride	2.788	84	8066	2.186	ug/l	98
43) Isopropyl Acetate	6.342	43	27251	2.656	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028464.D
Acq On : 30 Apr 2022 07:42
Operator : JC/MD
Sample : N2616-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-18D

Quant Time: Apr 30 09:25:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

