

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028465.D
 Acq On : 30 Apr 2022 08:05
 Operator : JC/MD
 Sample : N2616-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-20S

Quant Time: Apr 30 09:25:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	374649	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	672865	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	696078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	318465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	286989	58.749	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.500%
35) Dibromofluoromethane	5.385	113	254405	53.914	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.820%
50) Toluene-d8	8.653	98	966030	53.302	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.600%
62) 4-Bromofluorobenzene	11.079	95	355643	51.321	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.640%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	10220	5.484	ug/l	90
20) Methylene Chloride	2.794	84	18774	5.180	ug/l	95
43) Isopropyl Acetate	6.342	43	26757	2.632	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028465.D
Acq On : 30 Apr 2022 08:05
Operator : JC/MD
Sample : N2616-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-20S

Quant Time: Apr 30 09:25:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

