

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028438.D
 Acq On : 29 Apr 2022 20:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 09:18:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	113	0.00
2 T	Dichlorodifluoromethane	0.323	0.384	-18.9	151#	0.00
3 P	Chloromethane	0.375	0.384	-2.4	134	0.00
4 C	Vinyl Chloride	0.504	0.465	7.7#	117	0.00
5 T	Bromomethane	0.522	0.497	4.8	115	0.00
6 T	Chloroethane	0.461	0.325	29.5#	103	0.00
7 T	Trichlorofluoromethane	1.112	0.854	23.2	97	0.00
8 T	Diethyl Ether	0.382	0.326	14.7	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.405	0.444	-9.6	154#	0.00
10 T	Methyl Iodide	0.509	0.561	-10.2	129	0.00
11 T	Tert butyl alcohol	0.135	0.107	20.7	105	0.02
12 CM	1,1-Dichloroethene	0.384	0.410	-6.8#	149	0.00
13 T	Acrolein	0.080	0.073	8.8	96	0.00
14 T	Allyl chloride	0.645	0.583	9.6	111	0.00
15 T	Acrylonitrile	0.266	0.257	3.4	114	0.00
16 T	Acetone	0.215	0.208	3.3	128	0.00
17 T	Carbon Disulfide	1.035	0.817	21.1	100	0.00
18 T	Methyl Acetate	0.580	0.560	3.4	122	0.00
19 T	Methyl tert-butyl Ether	1.602	1.587	0.9	113	0.00
20 T	Methylene Chloride	0.484	0.500	-3.3	134	0.00
21 T	trans-1,2-Dichloroethene	0.523	0.474	9.4	108	0.00
22 T	Diisopropyl ether	1.469	1.390	5.4	112	0.00
23 T	Vinyl Acetate	1.283	1.204	6.2	108	0.00
24 P	1,1-Dichloroethane	0.867	0.837	3.5	120	0.00
25 T	2-Butanone	0.386	0.346	10.4	106	0.00
26 T	2,2-Dichloropropane	0.847	0.669	21.0	94	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.596	4.3	115	0.00
28 T	Bromochloromethane	0.372	0.334	10.2	103	0.00
29 T	Tetrahydrofuran	0.252	0.220	12.7	104	0.00
30 C	Chloroform	1.041	0.970	6.8#	110	0.00
31 T	Cyclohexane	0.794	0.673	15.2	100	0.00
32 T	1,1,1-Trichloroethane	0.943	0.852	9.7	105	0.00
33 S	1,2-Dichloroethane-d4	0.652	0.581	10.9	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	112	0.00
35 S	Dibromofluoromethane	0.351	0.332	5.4	111	0.00
36 T	1,1-Dichloropropene	0.457	0.407	10.9	106	0.00
37 T	Ethyl Acetate	0.464	0.424	8.6	108	0.00
38 T	Carbon Tetrachloride	0.516	0.460	10.9	102	0.00
39 T	Methylcyclohexane	0.577	0.461	20.1	94	0.00
40 TM	Benzene	1.339	1.245	7.0	108	0.00
41 T	Methacrylonitrile	0.239	0.220	7.9	109	0.00
42 TM	1,2-Dichloroethane	0.497	0.439	11.7	104	0.00
43 T	Isopropyl Acetate	0.755	0.671	11.1	106	0.00
44 TM	Trichloroethene	0.388	0.362	6.7	110	0.00
45 C	1,2-Dichloropropane	0.328	0.310	5.5#	113	0.00
46 T	Dibromomethane	0.257	0.241	6.2	112	0.00
47 T	Bromodichloromethane	0.494	0.448	9.3	104	0.00
48 T	Methyl methacrylate	0.347	0.318	8.4	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028438.D
 Acq On : 29 Apr 2022 20:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 09:18:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	99	0.00
50 S	Toluene-d8	1.347	1.197	11.1	103	0.00
51 T	4-Methyl-2-Pentanone	0.485	0.440	9.3	107	0.00
52 CM	Toluene	0.930	0.829	10.9#	103	0.00
53 T	t-1,3-Dichloropropene	0.525	0.471	10.3	100	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.509	9.8	104	0.00
55 T	1,1,2-Trichloroethane	0.378	0.362	4.2	112	0.00
56 T	Ethyl methacrylate	0.548	0.525	4.2	109	0.00
57 T	1,3-Dichloropropane	0.610	0.573	6.1	111	0.00
58 T	2-Chloroethyl Vinyl ether	0.257	0.249	3.1	106	0.00
59 T	2-Hexanone	0.375	0.341	9.1	106	0.00
60 T	Dibromochloromethane	0.412	0.379	8.0	102	0.00
61 T	1,2-Dibromoethane	0.415	0.383	7.7	105	0.00
62 S	4-Bromofluorobenzene	0.515	0.453	12.0	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.391	0.325	16.9	94	0.00
65 PM	Chlorobenzene	1.066	1.028	3.6	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.394	0.8	107	0.00
67 C	Ethyl Benzene	1.788	1.753	2.0#	103	0.00
68 T	m/p-Xylenes	0.728	0.702	3.6	100	0.00
69 T	o-Xylene	0.725	0.707	2.5	103	0.00
70 T	Styrene	1.175	1.178	-0.3	103	0.00
71 P	Bromoform	0.313	0.308	1.6	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.437	3.284	4.5	102	0.00
74 T	N-amyl acetate	1.188	1.228	-3.4	110	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	1.169	-4.8	114	0.00
76 T	1,2,3-Trichloropropane	0.984	0.982	0.2	104	0.00
77 T	Bromobenzene	0.895	0.863	3.6	105	0.00
78 T	n-propylbenzene	3.811	3.668	3.8	100	0.00
79 T	2-Chlorotoluene	2.369	2.250	5.0	101	0.00
80 T	1,3,5-Trimethylbenzene	2.917	2.762	5.3	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.331	0.301	9.1	93	0.00
82 T	4-Chlorotoluene	2.729	2.570	5.8	98	0.00
83 T	tert-Butylbenzene	2.979	2.879	3.4	101	0.00
84 T	1,2,4-Trimethylbenzene	2.900	2.773	4.4	97	0.00
85 T	sec-Butylbenzene	3.594	3.339	7.1	96	0.00
86 T	p-Isopropyltoluene	3.100	2.902	6.4	94	0.00
87 T	1,3-Dichlorobenzene	1.682	1.662	1.2	102	0.00
88 T	1,4-Dichlorobenzene	1.766	1.685	4.6	101	0.00
89 T	n-Butylbenzene	2.546	2.308	9.3	92	0.00
90 T	Hexachloroethane	0.498	0.475	4.6	92	0.00
91 T	1,2-Dichlorobenzene	1.671	1.680	-0.5	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.237	4.4	100	0.00
93 T	1,2,4-Trichlorobenzene	0.994	1.021	-2.7	109	0.00
94 T	Hexachlorobutadiene	0.443	0.416	6.1	103	0.00
95 T	Naphthalene	3.310	3.556	-7.4	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028438.D
Acq On : 29 Apr 2022 20:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 30 09:18:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.975	1.031	-5.7	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6