

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028441.D
 Acq On : 29 Apr 2022 22:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 09:19:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.323	0.401	-24.1	154#	0.00
3 P	Chloromethane	0.375	0.391	-4.3	132	0.00
4 C	Vinyl Chloride	0.504	0.483	4.2#	118	0.00
5 T	Bromomethane	0.522	0.471	9.8	106	0.00
6 T	Chloroethane	0.461	0.348	24.5	107	0.00
7 T	Trichlorofluoromethane	1.112	0.895	19.5	98	0.00
8 T	Diethyl Ether	0.382	0.338	11.5	111	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.405	0.460	-13.6	155#	0.00
10 T	Methyl Iodide	0.509	0.587	-15.3	131	0.00
11 T	Tert butyl alcohol	0.135	0.130	3.7	123	0.00
12 CM	1,1-Dichloroethene	0.384	0.426	-10.9#	150	0.00
13 T	Acrolein	0.080	0.076	5.0	96	0.00
14 T	Allyl chloride	0.645	0.609	5.6	113	0.00
15 T	Acrylonitrile	0.266	0.280	-5.3	120	0.00
16 T	Acetone	0.215	0.232	-7.9	139	0.00
17 T	Carbon Disulfide	1.035	0.837	19.1	100	0.00
18 T	Methyl Acetate	0.580	0.606	-4.5	129	0.00
19 T	Methyl tert-butyl Ether	1.602	1.645	-2.7	114	0.00
20 T	Methylene Chloride	0.484	0.508	-5.0	132	0.00
21 T	trans-1,2-Dichloroethene	0.523	0.496	5.2	109	0.00
22 T	Diisopropyl ether	1.469	1.422	3.2	111	0.00
23 T	Vinyl Acetate	1.283	1.245	3.0	108	0.00
24 P	1,1-Dichloroethane	0.867	0.859	0.9	120	0.00
25 T	2-Butanone	0.386	0.386	0.0	114	0.00
26 T	2,2-Dichloropropane	0.847	0.648	23.5	88	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.617	1.0	115	0.00
28 T	Bromochloromethane	0.372	0.334	10.2	100	0.00
29 T	Tetrahydrofuran	0.252	0.244	3.2	112	0.00
30 C	Chloroform	1.041	1.000	3.9#	110	0.00
31 T	Cyclohexane	0.794	0.699	12.0	101	0.00
32 T	1,1,1-Trichloroethane	0.943	0.879	6.8	105	0.00
33 S	1,2-Dichloroethane-d4	0.652	0.611	6.3	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.351	0.350	0.3	112	0.00
36 T	1,1-Dichloropropene	0.457	0.430	5.9	107	0.00
37 T	Ethyl Acetate	0.464	0.458	1.3	111	0.00
38 T	Carbon Tetrachloride	0.516	0.484	6.2	103	0.00
39 T	Methylcyclohexane	0.577	0.483	16.3	94	0.00
40 TM	Benzene	1.339	1.309	2.2	109	0.00
41 T	Methacrylonitrile	0.239	0.236	1.3	112	0.00
42 TM	1,2-Dichloroethane	0.497	0.464	6.6	105	0.00
43 T	Isopropyl Acetate	0.755	0.721	4.5	109	0.00
44 TM	Trichloroethene	0.388	0.391	-0.8	114	0.00
45 C	1,2-Dichloropropane	0.328	0.321	2.1#	111	0.00
46 T	Dibromomethane	0.257	0.251	2.3	111	0.00
47 T	Bromodichloromethane	0.494	0.472	4.5	105	0.00
48 T	Methyl methacrylate	0.347	0.344	0.9	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028441.D
 Acq On : 29 Apr 2022 22:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 09:19:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	112	0.00
50 S	Toluene-d8	1.347	1.281	4.9	105	0.00
51 T	4-Methyl-2-Pentanone	0.485	0.486	-0.2	113	0.00
52 CM	Toluene	0.930	0.885	4.8#	105	0.00
53 T	t-1,3-Dichloropropene	0.525	0.492	6.3	100	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.530	6.0	103	0.00
55 T	1,1,2-Trichloroethane	0.378	0.385	-1.9	114	0.00
56 T	Ethyl methacrylate	0.548	0.571	-4.2	113	0.00
57 T	1,3-Dichloropropane	0.610	0.608	0.3	113	0.00
58 T	2-Chloroethyl Vinyl ether	0.257	0.265	-3.1	108	0.00
59 T	2-Hexanone	0.375	0.380	-1.3	113	0.00
60 T	Dibromochloromethane	0.412	0.399	3.2	103	0.00
61 T	1,2-Dibromoethane	0.415	0.411	1.0	107	0.00
62 S	4-Bromofluorobenzene	0.515	0.494	4.1	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.391	0.393	-0.5	111	0.00
65 PM	Chlorobenzene	1.066	1.081	-1.4	108	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.416	-4.8	110	0.00
67 C	Ethyl Benzene	1.788	1.859	-4.0#	107	0.00
68 T	m/p-Xylenes	0.728	0.742	-1.9	104	0.00
69 T	o-Xylene	0.725	0.745	-2.8	106	0.00
70 T	Styrene	1.175	1.248	-6.2	107	0.00
71 P	Bromoform	0.313	0.325	-3.8	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.437	3.547	-3.2	106	0.00
74 T	N-amyl acetate	1.188	1.288	-8.4	110	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	1.211	-8.6	113	0.00
76 T	1,2,3-Trichloropropane	0.984	1.061	-7.8	108	0.00
77 T	Bromobenzene	0.895	0.936	-4.6	109	0.00
78 T	n-propylbenzene	3.811	3.969	-4.1	103	0.00
79 T	2-Chlorotoluene	2.369	2.409	-1.7	103	0.00
80 T	1,3,5-Trimethylbenzene	2.917	2.943	-0.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.331	0.316	4.5	93	0.00
82 T	4-Chlorotoluene	2.729	2.747	-0.7	101	0.00
83 T	tert-Butylbenzene	2.979	3.036	-1.9	102	0.00
84 T	1,2,4-Trimethylbenzene	2.900	2.979	-2.7	100	0.00
85 T	sec-Butylbenzene	3.594	3.582	0.3	98	0.00
86 T	p-Isopropyltoluene	3.100	3.090	0.3	96	0.00
87 T	1,3-Dichlorobenzene	1.682	1.763	-4.8	104	0.00
88 T	1,4-Dichlorobenzene	1.766	1.779	-0.7	103	0.00
89 T	n-Butylbenzene	2.546	2.495	2.0	95	0.00
90 T	Hexachloroethane	0.498	0.504	-1.2	94	0.00
91 T	1,2-Dichlorobenzene	1.671	1.767	-5.7	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.265	-6.9	107	0.00
93 T	1,2,4-Trichlorobenzene	0.994	1.073	-7.9	110	0.00
94 T	Hexachlorobutadiene	0.443	0.433	2.3	103	0.00
95 T	Naphthalene	3.310	3.828	-15.6	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
Data File : VX028441.D
Acq On : 29 Apr 2022 22:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 30 09:19:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.975	1.081	-10.9	111	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6