

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042922\  
 Data File : VX028423.D  
 Acq On : 29 Apr 2022 15:02  
 Operator : JC/MD  
 Sample : VX0429WBS02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0429WBS02

Quant Time: Apr 30 09:14:24 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 19 13:38:04 2022  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By : John Carlone 05/02/2022  
 Supervised By : Mahesh Dadoda 05/02/2022

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| -----                        |                |      |            |         |        |          |
| Internal Standards           |                |      |            |         |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 342837     | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 557504     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 529687     | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 274967     | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.958          | 65   | 202624     | 45.328  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = | 90.660% |        |          |
| 35) Dibromofluoromethane     | 5.391          | 113  | 177423     | 45.380  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = | 90.760% |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 658572     | 43.857  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = | 87.720% |        |          |
| 62) 4-Bromofluorobenzene     | 11.079         | 95   | 249150     | 43.393  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = | 86.780% |        |          |
| Target Compounds             |                |      |            |         |        |          |
|                              |                |      |            |         |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 50417      | 22.781  | ug/l   | 100      |
| 3) Chloromethane             | 1.294          | 50   | 49143      | 20.608  | ug/l   | 97       |
| 4) Vinyl Chloride            | 1.374          | 62   | 60143      | 17.401  | ug/l   | 99       |
| 5) Bromomethane              | 1.599          | 94   | 83111      | 22.486  | ug/l   | 97       |
| 6) Chloroethane              | 1.684          | 64   | 42451      | 14.887  | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 112984     | 14.824  | ug/l   | 96       |
| 8) Diethyl Ether             | 2.130          | 74   | 42607      | 16.520  | ug/l   | 90       |
| 9) 1,1,2-Trichlorotrifluo... | 2.331          | 101  | 57312      | 20.643  | ug/l   | 98       |
| 10) Methyl Iodide            | 2.453          | 142  | 67174      | 21.075  | ug/l   | 96       |
| 11) Tert butyl alcohol       | 2.959          | 59   | 74520      | 80.305  | ug/l   | 97       |
| 12) 1,1-Dichloroethene       | 2.318          | 96   | 52544      | 19.942  | ug/l   | 96       |
| 13) Acrolein                 | 2.233          | 56   | 52119      | 95.134  | ug/l   | 98       |
| 14) Allyl chloride           | 2.666          | 41   | 73580      | 16.643  | ug/l   | 91       |
| 15) Acrylonitrile            | 3.062          | 53   | 166010     | 91.130  | ug/l   | 99       |
| 16) Acetone                  | 2.373          | 43   | 144952     | 105.358 | ug/l   | 91       |
| 17) Carbon Disulfide         | 2.514          | 76   | 99801      | 14.066  | ug/l   | 100      |
| 18) Methyl Acetate           | 2.703          | 43   | 74168      | 18.661  | ug/l   | 98       |
| 19) Methyl tert-butyl Ether  | 3.111          | 73   | 196028     | 17.842  | ug/l   | 97       |
| 20) Methylene Chloride       | 2.788          | 84   | 63711      | 19.209  | ug/l   | 98       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 59452      | 16.592  | ug/l   | 93       |
| 22) Diisopropyl ether        | 3.757          | 45   | 173010     | 17.181  | ug/l # | 73       |
| 23) Vinyl Acetate            | 3.721          | 43   | 743267     | 84.488  | ug/l   | 100      |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 104535     | 17.582  | ug/l   | 98       |
| 25) 2-Butanone               | 4.556          | 43   | 226505     | 85.557  | ug/l   | 100      |
| 26) 2,2-Dichloropropane      | 4.483          | 77   | 85025      | 14.646  | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 4.489          | 96   | 71921      | 16.833  | ug/l   | 94       |
| 28) Bromochloromethane       | 4.897          | 49   | 50665      | 19.861  | ug/l   | 89       |
| 29) Tetrahydrofuran          | 5.001          | 42   | 143301     | 83.015  | ug/l   | 100      |
| 30) Chloroform               | 5.098          | 83   | 121237     | 16.979  | ug/l   | 96       |
| 31) Cyclohexane              | 5.476          | 56   | 84222      | 15.464  | ug/l   | 94       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 104916     | 16.226  | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 83492      | 16.371  | ug/l   | 98       |
| 37) Ethyl Acetate            | 4.714          | 43   | 85055      | 16.436  | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 5.678          | 117  | 87239      | 15.153  | ug/l   | 97       |
| 39) Methylcyclohexane        | 7.379          | 83   | 91815      | 14.268  | ug/l   | 95       |
| 40) Benzene                  | 6.043          | 78   | 249619     | 16.721  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042922\  
 Data File : VX028423.D  
 Acq On : 29 Apr 2022 15:02  
 Operator : JC/MD  
 Sample : VX0429WBS02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0429WBS02

Quant Time: Apr 30 09:14:24 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 19 13:38:04 2022  
 Response via : Initial Calibration

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 05/02/2022  
 Supervised By : Mahesh Dadoda 05/02/2022

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.922  | 41   | 45996    | 17.228  | ug/l   | 97       |
| 42) 1,2-Dichloroethane        | 6.092  | 62   | 92767    | 16.740  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 6.342  | 43   | 133886   | 15.897  | ug/l   | 98       |
| 44) Trichloroethene           | 7.129  | 130  | 72214    | 16.681  | ug/l   | 98       |
| 45) 1,2-Dichloropropane       | 7.433  | 63   | 61570    | 16.850  | ug/l   | 96       |
| 46) Dibromomethane            | 7.586  | 93   | 48497    | 16.919  | ug/l   | 98       |
| 47) Bromodichloromethane      | 7.824  | 83   | 86583    | 15.716  | ug/l   | 95       |
| 48) Methyl methacrylate       | 7.696  | 41   | 64286    | 16.626  | ug/l   | 92       |
| 49) 1,4-Dioxane               | 7.659  | 88   | 34133    | 329.979 | ug/l   | 94       |
| 51) 4-Methyl-2-Pentanone      | 8.573  | 43   | 465940   | 86.214  | ug/l   | 96       |
| 52) Toluene                   | 8.720  | 92   | 171749   | 16.560  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 86694    | 14.807  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 96373    | 15.328  | ug/l   | 92       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 71787    | 17.036  | ug/l   | 97       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 102601   | 16.797  | ug/l # | 91       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 114978   | 16.905  | ug/l   | 98       |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 259859   | 90.669  | ug/l   | 95       |
| 59) 2-Hexanone                | 9.433  | 43   | 359121   | 85.783  | ug/l   | 95       |
| 60) Dibromochloromethane      | 9.525  | 129  | 69026    | 15.026  | ug/l   | 96       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 76967    | 16.649  | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 69185    | 16.722  | ug/l   | 97       |
| 65) Chlorobenzene             | 10.079 | 112  | 194778   | 17.247  | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 70155    | 16.676  | ug/l   | 100      |
| 67) Ethyl Benzene             | 10.195 | 91   | 327044   | 17.270  | ug/l   | 97       |
| 68) m/p-Xylenes               | 10.305 | 106  | 266685   | 34.602  | ug/l   | 97       |
| 69) o-Xylene                  | 10.646 | 106  | 130414   | 16.972  | ug/l   | 97       |
| 70) Styrene                   | 10.658 | 104  | 215990   | 17.350  | ug/l   | 99       |
| 71) Bromoform                 | 10.799 | 173  | 50673    | 15.695  | ug/l # | 98       |
| 73) Isopropylbenzene          | 10.963 | 105  | 342155   | 18.104  | ug/l   | 100      |
| 74) N-amyl acetate            | 10.847 | 43   | 113672   | 17.401  | ug/l   | 97       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 115118   | 18.770  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 100266m  | 18.528  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 88981    | 18.078  | ug/l   | 91       |
| 78) n-propylbenzene           | 11.305 | 91   | 375127   | 17.900  | ug/l   | 97       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 232939   | 17.878  | ug/l   | 97       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 284733   | 17.748  | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 26227    | 14.403  | ug/l # | 82       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 264793   | 17.643  | ug/l   | 96       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 286901   | 17.513  | ug/l   | 97       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 285399   | 17.896  | ug/l   | 98       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 339373   | 17.171  | ug/l   | 98       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 293602   | 17.220  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 163975   | 17.730  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 167512   | 17.250  | ug/l   | 99       |
| 89) n-Butylbenzene            | 12.335 | 91   | 229679   | 16.405  | ug/l   | 97       |
| 90) Hexachloroethane          | 12.542 | 117  | 42322    | 15.108  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 166109   | 18.080  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 22543    | 16.528  | ug/l   | 93       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 96438    | 17.648  | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 41355    | 16.960  | ug/l   | 93       |
| 95) Naphthalene               | 13.780 | 128  | 341777   | 18.775  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 95391    | 17.784  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042922\  
Data File : VX028423.D  
Acq On : 29 Apr 2022 15:02  
Operator : JC/MD  
Sample : VX0429WBS02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02

Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022

Quant Time: Apr 30 09:14:24 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 19 13:38:04 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 71  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion: 168 Resp: 34283  
Ion Ratio Lower Upper  
168 100  
99 57.1 41.7 62.5




Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#2  
Dichlorodifluoromethane  
Concen: 22.781 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02

Tgt Ion: 85 Resp: 50417  
Ion Ratio Lower Upper  
85 100  
87 32.6 16.2 48.6




Abundance Scan 34 (1.294 min): VX028153.D\data.ms (-28)



#3  
Chloromethane  
Concen: 20.608 ug/l  
RT: 1.294 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion: 50 Resp: 4914  
Ion Ratio Lower Upper  
50 100  
52 34.5 26.4 39.6




Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#4  
Vinyl Chloride  
Concen: 17.401 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion: 62 Resp: 60143  
Ion Ratio Lower Upper  
62 100  
64 32.5 25.5 38.3





#5  
Bromomethane  
Concen: 22.486 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 94 Resp: 8311  
Ion Ratio Lower Upper  
94 100  
96 97.3 75.3 112.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#6  
Chloroethane  
Concen: 14.887 ug/l  
RT: 1.684 min Scan# 98  
Delta R.T. 0.006 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 64 Resp: 42451  
Ion Ratio Lower Upper  
64 100  
66 31.1 25.3 37.9





#7  
Trichlorofluoromethane  
Concen: 14.824 ug/l  
RT: 1.886 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:101 Resp: 112984  
Ion Ratio Lower Upper  
101 100  
103 66.4 50.5 75.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#8  
Diethyl Ether  
Concen: 16.520 ug/l  
RT: 2.130 min Scan# 171  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 74 Resp: 42607  
Ion Ratio Lower Upper  
74 100  
45 75.3 42.0 126.0





#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 20.643 ug/l  
RT: 2.331 min Scan# 204  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:101 Resp: 57312  
Ion Ratio Lower Upper  
101 100  
85 43.1 33.5 50.3  
151 78.5 61.8 92.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#10  
Methyl Iodide  
Concen: 21.075 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Tgt Ion:142 Resp: 67174  
Ion Ratio Lower Upper  
142 100  
127 45.8 33.9 50.9  
141 14.2 11.4 17.2





#11

Tert butyl alcohol

Concen: 80.305 ug/l

RT: 2.959 min Scan# 308

Delta R.T. -0.006 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02

Tgt Ion: 59 Resp: 74520

Ion Ratio Lower Upper

59 100





57 9.5 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#12

1,1-Dichloroethene

Concen: 19.942 ug/l

RT: 2.318 min Scan# 202

Delta R.T. -0.001 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02




Tgt Ion: 96 Resp: 52544


Ion Ratio Lower Upper

96 100

61 148.6 124.6 187.0


98 64.5 51.0 76.4





#13  
Acrolein  
Concen: 95.134 ug/l  
RT: 2.233 min Scan# 188  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 56 Resp: 52119  
Ion Ratio Lower Upper  
56 100  
55 70.4 55.0 82.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#14  
Allyl chloride  
Concen: 16.643 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 41 Resp: 73580  
Ion Ratio Lower Upper  
41 100  
39 73.8 50.4 75.6  
76 39.4 32.7 49.1





#15  
Acrylonitrile  
Concen: 91.130 ug/l  
RT: 3.062 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion: 53 Resp: 166010  
Ion Ratio Lower Upper  
53 100  
52 82.4 65.7 98.5  
51 36.2 28.4 42.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#16  
Acetone  
Concen: 105.358 ug/l  
RT: 2.373 min Scan# 211  
Delta R.T. -0.007 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion: 43 Resp: 144952  
Ion Ratio Lower Upper  
43 100  
58 30.9 28.8 43.2





#17  
Carbon Disulfide  
Concen: 14.066 ug/l  
RT: 2.514 min Scan# 21  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 76 Resp: 9980  
Ion Ratio Lower Upper  
76 100  
78 9.0 7.2 10.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#18  
Methyl Acetate  
Concen: 18.661 ug/l  
RT: 2.703 min Scan# 265  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 43 Resp: 74168  
Ion Ratio Lower Upper  
43 100  
74 27.5 22.8 34.2





#19  
Methyl tert-butyl Ether  
Concen: 17.842 ug/l  
RT: 3.111 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 73 Resp: 196028  
Ion Ratio Lower Upper  
73 100  
57 19.4 16.8 25.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#20  
Methylene Chloride  
Concen: 19.209 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 84 Resp: 63711  
Ion Ratio Lower Upper  
84 100  
49 106.0 85.5 128.3  
51 34.6 26.6 39.8  
86 67.3 51.5 77.3





#21  
trans-1,2-Dichloroethene  
Concen: 16.592 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 96 Resp: 59452  
Ion Ratio Lower Upper  
96 100  
61 121.8 105.4 158.2  
98 60.0 50.0 75.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#22  
Diisopropyl ether  
Concen: 17.181 ug/l  
RT: 3.757 min Scan# 438  
Delta R.T. -0.006 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 45 Resp: 173010  
Ion Ratio Lower Upper  
45 100  
43 98.8 50.8 76.2#  
87 34.7 26.6 39.8  
59 13.2 10.5 15.7





#23  
 Vinyl Acetate  
 Concen: 84.488 ug/l  
 RT: 3.721 min Scan# 41  
 Delta R.T. -0.000 min  
 Lab File: VX028423.D  
 Acq: 29 Apr 2022 15:02




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0429WBS02




Tgt Ion: 43 Resp: 74326  
 Ion Ratio Lower Upper  
 43 100  
 86 13.7 10.8 16.2

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 05/02/2022  
 Supervised By :Mahesh Dadoda 05/02/2022




#24  
 1,1-Dichloroethane  
 Concen: 17.582 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. -0.000 min  
 Lab File: VX028423.D  
 Acq: 29 Apr 2022 15:02




Tgt Ion: 63 Resp: 104535  
 Ion Ratio Lower Upper  
 63 100  
 98 7.4 3.9 11.7  
 100 5.5 2.4 7.2





#25  
2-Butanone  
Concen: 85.557 ug/l  
RT: 4.556 min Scan# 569  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 43 Resp: 22650  
Ion Ratio Lower Upper  
43 100  
72 30.4 24.4 36.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#26  
2,2-Dichloropropane  
Concen: 14.646 ug/l  
RT: 4.483 min Scan# 557  
Delta R.T. 0.006 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 77 Resp: 85025  
Ion Ratio Lower Upper  
77 100  
97 24.7 11.8 35.4





#27  
cis-1,2-Dichloroethene  
Concen: 16.833 ug/l  
RT: 4.489 min Scan# 511  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 96 Resp: 7192  
Ion Ratio Lower Upper  
96 100  
61 131.9 0.0 283.8  
98 65.9 0.0 129.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#28  
Bromochloromethane  
Concen: 19.861 ug/l  
RT: 4.897 min Scan# 625  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 49 Resp: 50665  
Ion Ratio Lower Upper  
49 100  
129 2.8 0.0 4.4  
130 98.7 70.7 106.1





#29  
Tetrahydrofuran  
Concen: 83.015 ug/l  
RT: 5.001 min Scan# 643  
Delta R.T. -0.006 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion: 42 Resp: 14330  
Ion Ratio Lower Upper  
42 100  
72 55.4 44.1 66.1  
71 50.7 40.5 60.7




Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#30  
Chloroform  
Concen: 16.979 ug/l  
RT: 5.098 min Scan# 658  
Delta R.T. -0.001 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion: 83 Resp: 121237  
Ion Ratio Lower Upper  
83 100  
85 63.0 52.7 79.1





#31  
Cyclohexane  
Concen: 15.464 ug/l  
RT: 5.476 min Scan# 719  
Delta R.T. 0.006 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion: 56 Resp: 8422  
Ion Ratio Lower Upper  
56 100  
69 32.0 26.7 40.1  
84 103.5 77.2 115.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#32  
1,1,1-Trichloroethane  
Concen: 16.226 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion: 97 Resp: 104916  
Ion Ratio Lower Upper  
97 100  
99 62.6 51.8 77.6  
61 40.9 34.0 51.0





#33  
1,2-Dichloroethane-d4  
Concen: 45.328 ug/l  
RT: 5.958 min Scan# 799  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion: 65 Resp: 202624  
Ion Ratio Lower Upper  
65 100  
67 51.9 0.0 107.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.763 min Scan# 931  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion:114 Resp: 557504  
Ion Ratio Lower Upper  
114 100  
63 18.7 0.0 36.8  
88 15.7 0.0 32.4





#35  
Dibromofluoromethane  
Concen: 45.380 ug/l  
RT: 5.391 min Scan# 705  
Delta R.T. 0.006 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 113 Resp: 17742  
Ion Ratio Lower Upper  
113 100  
111 103.2 82.8 124.2  
192 19.9 16.9 25.3

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#36  
1,1-Dichloropropene  
Concen: 16.371 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Tgt Ion: 75 Resp: 83492  
Ion Ratio Lower Upper  
75 100  
110 38.0 18.3 54.8  
77 30.8 24.8 37.2





#37

Ethyl Acetate

Concen: 16.436 ug/l

RT: 4.714 min Scan# 594

Delta R.T. 0.006 min

Lab File: VX028423.D


Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02





|             |                |
|-------------|----------------|
| Tgt Ion: 43 | Resp: 8505     |
| Ion Ratio   | Lower Upper    |
| 43          | 100            |
| 61          | 15.6 12.6 19.0 |
| 70          | 13.2 10.2 15.2 |

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#38

Carbon Tetrachloride

Concen: 15.153 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.006 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

|             |                 |
|-------------|-----------------|
| Tgt Ion:117 | Resp: 87239     |
| Ion Ratio   | Lower Upper     |
| 117         | 100             |
| 119         | 93.7 77.3 115.9 |
| 121         | 29.4 24.2 36.2  |





#39

Methylcyclohexane

Concen: 14.268 ug/l

RT: 7.379 min Scan# 1033

Delta R.T. -0.006 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02

Tgt Ion: 83 Resp: 9181





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 55  | 65.6  | 54.6  | 81.8  |
| 98  | 52.1  | 37.4  | 56.2  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#40

Benzene

Concen: 16.721 ug/l

RT: 6.043 min Scan# 813




Delta R.T. -0.001 min


Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Tgt Ion: 78 Resp: 249619


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 78  | 100   |       |       |
| 77  | 24.1  | 19.3  | 28.9  |





#41  
Methacrylonitrile  
Concen: 17.228 ug/l  
RT: 4.922 min Scan# 61  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 41 Resp: 45990  
Ion Ratio Lower Upper  
41 100  
39 56.4 43.0 64.4  
67 85.9 70.2 105.4  
52 33.8 26.0 39.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#42  
1,2-Dichloroethane  
Concen: 16.740 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 62 Resp: 92767  
Ion Ratio Lower Upper  
62 100  
98 9.8 0.0 20.4





#43  
Isopropyl Acetate  
Concen: 15.897 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. 0.006 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion: 43 Resp: 133880  
Ion Ratio Lower Upper  
43 100  
61 21.3 18.5 27.7  
87 16.4 13.0 19.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#44  
Trichloroethene  
Concen: 16.681 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02

Tgt Ion:130 Resp: 72214  
Ion Ratio Lower Upper  
130 100  
95 96.7 0.0 188.8





#45

1,2-Dichloropropane

Concen: 16.850 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.001 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02

Tgt Ion: 63 Resp: 61570

Ion Ratio Lower Upper

63 100





65 29.9 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#46

Dibromomethane

Concen: 16.919 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.006 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02




Tgt Ion: 93 Resp: 48497


Ion Ratio Lower Upper

93 100

95 82.4 65.8 98.6

174 102.7 85.4 128.2





#47

Bromodichloromethane

Concen: 15.716 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02

Tgt Ion: 83 Resp: 8658





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 85  | 60.6  | 52.2  | 78.2  |
| 127 | 7.9   | 6.8   | 10.2  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#48

Methyl methacrylate

Concen: 16.626 ug/l

RT: 7.696 min Scan# 1084




Delta R.T. -0.000 min


Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Tgt Ion: 41 Resp: 64286

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 41  | 100   |       |       |
| 69  | 98.4  | 82.5  | 123.7 |
| 39  | 64.5  | 43.8  | 65.8  |





#49

1,4-Dioxane

Concen: 329.979 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02

Tgt Ion: 88 Resp: 3413




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 88  | 100   |       |       |
| 43  | 29.1  | 21.9  | 32.9  |
| 58  | 61.0  | 53.7  | 80.5  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#50

Toluene-d8

Concen: 43.857 ug/l

RT: 8.653 min Scan# 1241




Delta R.T. -0.000 min


Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Tgt Ion: 98 Resp: 658572


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 98  | 100   |       |       |
| 100 | 66.7  | 52.6  | 78.8  |





#51  
4-Methyl-2-Pentanone  
Concen: 86.214 ug/l  
RT: 8.573 min Scan# 1228  
Delta R.T. -0.001 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 43 Resp: 465940  
Ion Ratio Lower Upper  
43 100  
58 42.5 36.1 54.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#52  
Toluene  
Concen: 16.560 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Tgt Ion: 92 Resp: 171749  
Ion Ratio Lower Upper  
92 100  
91 170.6 137.5 206.3





#53

t-1,3-Dichloropropene

Concen: 14.807 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02

Tgt Ion: 75 Resp: 86694

Ion Ratio Lower Upper

75 100





77 31.1 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#54

cis-1,3-Dichloropropene

Concen: 15.328 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02




Tgt Ion: 75 Resp: 96373


Ion Ratio Lower Upper

75 100

77 32.4 24.8 37.2


39 44.3 29.6 44.4





#55  
1,1,2-Trichloroethane  
Concen: 17.036 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02






Tgt Ion: 97 Resp: 7178


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 84.6  | 66.2  | 99.4  |
| 85  | 54.1  | 42.9  | 64.3  |
| 99  | 65.1  | 49.0  | 73.6  |

Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022





#56  
Ethyl methacrylate  
Concen: 16.797 ug/l  
RT: 9.116 min Scan# 1317  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Tgt Ion: 69 Resp: 102601


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 57.3  | 43.8  | 65.6  |
| 39  | 37.4  | 22.2  | 33.2  |





#57  
1,3-Dichloropropane  
Concen: 16.905 ug/l  
RT: 9.311 min Scan# 114978  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 76 Resp: 114978  
Ion Ratio Lower Upper  
76 100  
78 33.6 25.8 38.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#58  
2-Chloroethyl Vinyl ether  
Concen: 90.669 ug/l  
RT: 8.244 min Scan# 1174  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 63 Resp: 259859  
Ion Ratio Lower Upper  
63 100  
106 32.8 23.9 35.9





#59  
2-Hexanone  
Concen: 85.783 ug/l  
RT: 9.433 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion: 43 Resp: 35912  
Ion Ratio Lower Upper  
43 100  
58 60.0 31.9 95.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#60  
Dibromochloromethane  
Concen: 15.026 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion:129 Resp: 69026  
Ion Ratio Lower Upper  
129 100  
127 77.2 40.3 120.8





#61  
1,2-Dibromoethane  
Concen: 16.649 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:107 Resp: 7696  
Ion Ratio Lower Upper  
107 100  
109 94.4 74.4 111.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#62  
4-Bromofluorobenzene  
Concen: 43.393 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 95 Resp: 249150  
Ion Ratio Lower Upper  
95 100  
174 85.9 0.0 163.0  
176 84.7 0.0 155.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion:117 Resp: 52968  
Ion Ratio Lower Upper  
117 100  
82 52.9 41.8 62.8  
119 31.4 24.8 37.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#64  
Tetrachloroethene  
Concen: 16.722 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion:164 Resp: 69185  
Ion Ratio Lower Upper  
164 100  
166 121.3 101.2 151.8  
129 89.7 73.0 109.4  
131 90.8 74.2 111.2





#65  
Chlorobenzene  
Concen: 17.247 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:112 Resp: 194778  
Ion Ratio Lower Upper  
112 100  
114 31.4 25.8 38.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 16.676 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Tgt Ion:131 Resp: 70155  
Ion Ratio Lower Upper  
131 100  
133 97.0 48.3 144.8  
119 64.9 32.6 98.0





#67

Ethyl Benzene

Concen: 17.270 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02

Tgt Ion: 91 Resp: 327044

Ion Ratio Lower Upper

91 100





106 32.1 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#68

m/p-Xylenes

Concen: 34.602 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. -0.000 min

Lab File: VX028423.D




Acq: 29 Apr 2022 15:02


Tgt Ion:106 Resp: 266685

Ion Ratio Lower Upper

106 100


91 195.5 160.4 240.6





#69  
o-Xylene  
Concen: 16.972 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02



Tgt Ion:106 Resp: 130414  
Ion Ratio Lower Upper  
106 100  
91 207.7 106.1 318.1




Manual Integrations  
APPROVED


Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#70  
Styrene  
Concen: 17.350 ug/l  
RT: 10.658 min Scan# 1570  
Delta R.T. -0.001 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion:104 Resp: 215990  
Ion Ratio Lower Upper  
104 100  
78 46.8 38.6 57.8  
103 55.2 43.8 65.8





#71  
Bromoform  
Concen: 15.695 ug/l  
RT: 10.799 min Scan# 1593  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:173 Resp: 5067  
Ion Ratio Lower Upper  
173 100  
175 49.1 24.1 72.2  
254 0.0 0.1 0.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion:152 Resp: 274967  
Ion Ratio Lower Upper  
152 100  
115 65.5 37.5 112.7  
150 163.3 0.0 338.2





#73  
Isopropylbenzene  
Concen: 18.104 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:105 Resp: 34215  
Ion Ratio Lower Upper  
105 100  
120 26.9 13.5 40.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#74  
N-ethyl acetate  
Concen: 17.401 ug/l  
RT: 10.847 min Scan# 1601  
Delta R.T. -0.001 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 43 Resp: 113672  
Ion Ratio Lower Upper  
43 100  
70 50.9 42.5 63.7  
55 26.9 23.1 34.7  
61 26.8 22.8 34.2





#75  
1,1,2,2-Tetrachloroethane  
Concen: 18.770 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 83 Resp: 115118  
Ion Ratio Lower Upper  
83 100  
131 10.7 5.1 15.3  
85 63.7 32.6 97.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#76  
1,2,3-Trichloropropane  
Concen: 18.528 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 75 Resp: 100266  
Ion Ratio Lower Upper  
75 100  
77 37.3 19.4 58.2





#77  
Bromobenzene  
Concen: 18.078 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:156 Resp: 8898  
Ion Ratio Lower Upper  
156 100  
77 141.7 78.5 235.3  
158 94.8 50.2 150.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#78  
n-propylbenzene  
Concen: 17.900 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Tgt Ion: 91 Resp: 375127  
Ion Ratio Lower Upper  
91 100  
120 24.8 11.6 34.8









#79  
2-Chlorotoluene  
Concen: 17.878 ug/l  
RT: 11.366 min Scan# 1100  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02

Tgt Ion: 91 Resp: 232939  
Ion Ratio Lower Upper  
91 100  
126 36.5 17.3 51.9




Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#80  
1,3,5-Trimethylbenzene  
Concen: 17.748 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02


Tgt Ion:105 Resp: 284733  
Ion Ratio Lower Upper  
105 100  
120 50.1 24.3 72.9





#81  
trans-1,4-Dichloro-2-butene  
Concen: 14.403 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 75 Resp: 26227  
Ion Ratio Lower Upper  
75 100  
53 108.3 69.0 103.4  
89 49.8 37.0 55.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#82  
4-Chlorotoluene  
Concen: 17.643 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 91 Resp: 264793  
Ion Ratio Lower Upper  
91 100  
126 32.0 14.9 44.5





#83  
tert-Butylbenzene  
Concen: 17.513 ug/l  
RT: 11.719 min Scan# 11744  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:119 Resp: 28690  
Ion Ratio Lower Upper  
119 100  
91 55.5 26.9 80.6  
134 24.1 11.4 34.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#84  
1,2,4-Trimethylbenzene  
Concen: 17.896 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion:105 Resp: 285399  
Ion Ratio Lower Upper  
105 100  
120 46.8 22.6 67.8





#85  
 sec-Butylbenzene  
 Concen: 17.171 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. -0.000 min  
 Lab File: VX028423.D  
 Acq: 29 Apr 2022 15:02




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0429WBS02




Tgt Ion:105 Resp: 33937  
 Ion Ratio Lower Upper  
 105 100  
 134 21.2 10.2 30.4

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 05/02/2022  
 Supervised By :Mahesh Dadoda 05/02/2022




#86  
 p-Isopropyltoluene  
 Concen: 17.220 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. -0.000 min  
 Lab File: VX028423.D  
 Acq: 29 Apr 2022 15:02



Tgt Ion:119 Resp: 293602  
 Ion Ratio Lower Upper  
 119 100  
 134 27.0 13.0 39.0  
 91 22.5 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 17.730 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02

Instrument :

MSVOA\_X

ClientSampleId :

VX0429WBS02





|             |       |           |
|-------------|-------|-----------|
| Tgt Ion:146 | Resp: | 16397     |
| Ion Ratio   | Lower | Upper     |
| 146         | 100   |           |
| 111         | 38.8  | 19.7 59.1 |
| 148         | 64.4  | 31.6 94.7 |

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/02/2022

Supervised By :Mahesh Dadoda 05/02/2022



#88

1,4-Dichlorobenzene

Concen: 17.250 ug/l




RT: 12.042 min Scan# 1797


Delta R.T. 0.000 min

Lab File: VX028423.D

Acq: 29 Apr 2022 15:02


|             |       |           |
|-------------|-------|-----------|
| Tgt Ion:146 | Resp: | 167512    |
| Ion Ratio   | Lower | Upper     |
| 146         | 100   |           |
| 111         | 39.1  | 19.2 57.6 |
| 148         | 65.0  | 32.3 96.9 |





#89  
n-Butylbenzene  
Concen: 16.405 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion: 91 Resp: 229679  
Ion Ratio Lower Upper  
91 100  
92 54.0 27.8 83.3  
134 28.4 13.3 39.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#90  
Hexachloroethane  
Concen: 15.108 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion:117 Resp: 42322  
Ion Ratio Lower Upper  
117 100  
201 78.4 39.5 118.5





#91  
1,2-Dichlorobenzene  
Concen: 18.080 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:146 Resp: 166109  
Ion Ratio Lower Upper  
146 100  
111 39.6 21.0 63.0  
148 64.4 32.6 97.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 16.528 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion: 75 Resp: 22543  
Ion Ratio Lower Upper  
75 100  
155 100.6 47.0 141.2  
157 124.4 57.9 173.6





#93  
1,2,4-Trichlorobenzene  
Concen: 17.648 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:180 Resp: 96438  
Ion Ratio Lower Upper  
180 100  
182 96.7 47.7 143.1  
145 31.5 15.2 45.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022




#94  
Hexachlorobutadiene  
Concen: 16.960 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Tgt Ion:225 Resp: 41355  
Ion Ratio Lower Upper  
225 100  
223 60.6 34.2 102.5  
227 63.4 33.9 101.7





#95  
Naphthalene  
Concen: 18.775 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0429WBS02




Tgt Ion:128 Resp: 34177  
Ion Ratio Lower Upper  
128 100  
127 12.8 10.2 15.4  
129 10.8 8.7 13.1

Manual Integrations  
APPROVED

Reviewed By :John Carlone 05/02/2022  
Supervised By :Mahesh Dadoda 05/02/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 17.784 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. -0.000 min  
Lab File: VX028423.D  
Acq: 29 Apr 2022 15:02



Tgt Ion:180 Resp: 95391  
Ion Ratio Lower Upper  
180 100  
182 96.1 50.0 149.8  
145 32.5 16.2 48.5

