


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042922\
 Data File : VX028440.D
 Acq On : 29 Apr 2022 22:01
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Title : SW846 8260
 Last Update : Tue Apr 19 13:38:04 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	14199	PASS
75	95	30	60	52.4	41595	PASS
95	95	100	100	100.0	79360	PASS
96	95	5	9	6.8	5433	PASS
173	174	0.00	2	1.2	741	PASS
174	95	50	100	78.6	62416	PASS
175	174	5	9	7.4	4591	PASS
176	174	95	101	96.0	59923	PASS
177	176	5	9	6.9	4144	PASS